

# VERIFICATION OF CALIBRATION REPORT

*of DataMaster cdm Breath Test Instrument  
State of Alaska*

*Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program*

DataMaster cdm S/N 130185 ✓

Supervisor/Operator Performing the Verification Procedure:  
 Name TODD BRUND ID# 7672 ✓ Date 11/09/10 ✓

A Agency NPS GLACIER BAY NP Phone # 907 697-2687  
 Instrument Location BARTLETT COVE RANGER STATION

---

B Alco S/N 2K ✓ Target Value 0.080 ✓ High Pressure 1000 psi

Alco Test Values 0.078 ✓ 0.079 ✓  
 1st Alco 2<sup>nd</sup> Alco

Signature Todd Brund

(OVER)

*BMB  
12/16/10*

(Do Not write in the area below)

I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

*Nita J. Bolz*  
 Nita J. Bolz  
 Scientific Director  
 State Breath Alcohol Program

Subscribed and sworn before me this 23<sup>rd</sup> day of DEC, 2010.

*Daniel L. Nicholas*  
 (Notary Seal Stamp)  
 Daniel L. Nicholas  
 Notary Public, State of Alaska  
 Commission Expires with Office



