

# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N <u>130185</u>	
Supervisor/Operator Performing the Verification Procedure:	
Name <u>TODD BRUND</u>	ID# <u>7672</u> Date <u>11/15/11</u>
A Agency <u>GRANGER BAY NATIONAL PARK</u>	Phone # <u>907 697 2687</u>
Instrument Location <u>BARTLETT CAVE, AK</u>	
B Alco S/N <u>2K</u> Target Value <u>.080</u> High Pressure <u>700 psi</u>	
Alco Test Values <u>0.081</u> <u>0.081</u>	
1st Alco 2nd Alco	
Signature <u>Todd Brund</u>	

(OVER)

BMB  
11/22/11

(Do Not write in the area below)


I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

Nita J. Bolz  
Nita J. Bolz  
Scientific Director  
State Breath Alcohol Program

Subscribed and sworn before me this 1st day of Feb 2012 me

Nikki Roth (Notary Seal Stamp)  
~~me~~ Daniel L. Nicholas  
Notary Public, State of Alaska  
Commission Expires with Office



# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130185

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

## NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130185

NOVEMBER 15, 2011

OPERATOR'S NAME:

BRUND/TODD/W

OPERATOR'S NUMBER: 7572

SUBJECT'S LAST NAME:

TEST

SUBJECT'S FIRST NAME/MI:

TEST

O.L. #: 0000

DEPT/AGENCY: 6030

ORIG/REPORT: 0040

TEST TYPE: T

ALCO TARGET VALUE: 0.150

ALCO S/N: 2K

### --- BREATH ANALYSIS ---

.000 ADJUSTED FOR 30.16 in  
ALCO TARGET .000 13:33  
BLANK TEST .000 13:34  
INTERNAL STANDARD VERIFIED 2.134  
ALCO TV 30.16 in .001 13:34  
BLANK TEST .000 13:35  
SUBJECT SAMPLE .000 13:35  
BLANK TEST .000 13:36  
ALCO TV 30.16 in .001 13:37  
BLANK TEST .000 13:37

## DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130185

NOVEMBER 15, 2011

TIME 13:30

### --- DIAGNOSTIC CHECK ---

COMPUTER: OKAY  
PROGRAM: OKAY  
SOFTWARE DATE: 01/30/09  
HEATERS  
SAMPLE CHAMBER: 43c  
BREATH TUBIC: 42c  
BAROMETER: 30.16 in  
FLOW DETECTOR: OKAY  
PUMP  
HIGH SPEED: OKAY  
DETECTOR: OKAY  
FILTERS: OKAY  
QUARTZ STANDARD: OKAY  
CALIBRATION: OKAY

### PRINTER TEST

ALCO TARGET VALUE: 0.150  
ALCO TV 30.16 in: 0.001  
SUBJECT SAMPLE: 0.000

