

# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Supervisor/Operator Performing the Verification Procedure: DataMaster cdm S/N 130186

Name Bill J. Ferris ID# 5004 Date 11-29-07

A Agency Togot Police Department Phone # \_\_\_\_\_

Instrument Location 100 Main St. Togot (Police station)

B Alco S/N X301540 Target Value .083 High Pressure 1000 psi

Alco Test Value Average 0.83 0.83  
1<sup>st</sup> Alco 2<sup>nd</sup> Alco

Signature Bill J Ferris CB  
3/10/08

(OVER)

(Do Not write in the area below)


Under the Alaska Rules of Evidence, I certify that;

(a) I am Jeanne Swartz, Criminalist III, Scientific Crime Detection Laboratory, State of Alaska Department of Public Safety, and was, at the time of signing, the custodian of the records of the aforementioned laboratory and the signature was made in my official capacity as Scientific Director Designee of the State Breath Alcohol Testing Program.

(b) The verification of calibration for the above instrument has been received and reviewed. Instrument operator, Bill J Ferris, whose name appears in my records, is qualified to verify the calibration of the above instrument.

(c) This instrument is certified for evidentiary use and these records are being retained as original documents in my custody.

Signed and sealed,



BT9 03/00

Jeanne Swartz

Jeanne Swartz  
Scientific Director Designee  
State of Alaska  
State Breath Alcohol Testing Program

03/10/08

Date

