

VERIFICATION OF CALIBRATION REPORT

SEP 14 2011

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

| | | |
|--|-------------------------|--|
| Supervisor/Operator Performing the Verification Procedure: | | DataMaster cdm S/N <u>130187</u> |
| Name | <u>DAN PASQUARIELLO</u> | ID# <u>2620</u> Date <u>9-10-11</u> |
| A Agency | <u>DILLINGHAM PD</u> | Phone # <u>847-5354</u> |
| Instrument Location | <u>DILLINGHAM PD</u> | |
| B Alco S/N | <u>58428</u> | Target Value <u>.080</u> High Pressure <u>1100 PSI</u> |
| Alco Test Values | <u>.080</u> 1st Alco | <u>.080</u> 2nd Alco |
| Signature | <u>[Signature]</u> | <u>CSB</u> <u>9/15/11</u> |

(OVER)

(Do Not write in the area below)

I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

[Signature]
Nita J. Bolz
Scientific Director
State Breath Alcohol Program

Subscribed and sworn before me this 9 day of Nov, 2011.

[Signature] (Notary Seal Stamp)
Daniel L. Nicholas
Notary Public, State of Alaska
Commission Expires with Office



VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument
State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N

130187

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130187

SEPTEMBER 10, 2011

OPERATOR'S NAME:

DAN PASQUARIELLO

OPERATOR'S NUMBER: 2620

SUBJECT'S LAST NAME:

VOC

SUBJECT'S FIRST NAME/MI :

VOC

D.L. #: 1234567890

DEPT/AGENCY: DL61

CASE/REPORT: 1234567890

TEST TYPE: V

ALCO TARGET VALUE: .000

ALCO S/N: 58428

--- BREATH ANALYSIS ---

| | | |
|----------------------------|----------|-------|
| .000 ADJUSTED FOR 30.32 in | | |
| ALCO TARGET | .001 | 11:06 |
| BLANK TEST | .000 | 11:07 |
| INTERNAL STANDARD | VERIFIED | 11:07 |
| ALCO TV 30.32 in | .000 | 11:07 |
| BLANK TEST | .000 | 11:08 |
| SUBJECT SAMPLE | .000 | 11:09 |
| BLANK TEST | .000 | 11:09 |
| ALCO TV 30.32 in | .000 | 11:10 |
| BLANK TEST | .000 | 11:11 |

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130187

SEPTEMBER 10, 2011

TIME 11:03

--- DIAGNOSTIC CHECK ---

| | |
|------------------|----------|
| COMPUTER: | OKAY |
| PROGRAM: | OKAY |
| SOFTWARE DATE: | 02/20/01 |
| HEATERS | |
| SAMPLE CHAMBER: | 42c |
| BREATH TUBE: | 42c |
| BAROMETER: | 30.32 in |
| FLOW DETECTOR: | OKAY |
| PUMP | |
| HIGH SPEED: | OKAY |
| DETECTOR: | OKAY |
| FILTERS: | OKAY |
| QUARTZ STANDARD: | OKAY |
| CALIBRATION: | OKAY |

PRINTER TEST

! " # \$ % & ' () * + , - . / 0 1 2 3 4 5 6 7 8 9 : ; (=) ? @ A B C D E F G
H I J K L M N O P Q R S T U V W X Y Z [\] ^ _ ` a b c d e f g h i j k l m n o
p q r s t u v w x y z { | } ~ ¡ ¢

