VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument
State of Alaska
Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130191

Supervisor/Operator Performing the Verification Procedure:

A

Name Robert J. Ely ID# 2681 Date 01-22-2008

Agency Craig Police Department Phone # (907) 826-3330

Instrument Location 506 2nd Street, Craig Alaska 99921

Alco S/N X172967 Target Value .081 High Pressure 600 PSI

Alco Test Value Average .086 (1st Alco) .087 (2nd Alco)

Signature Robert J. Ely CDB 2/19/08

(OVER)

(Do Not write in the area below)

Under the Alaska Rules of Evidence, I certify that;

(a) I am Jeanne Swartz , Criminalist III, Scientific Crime Detection Laboratory, State of Alaska Department of Public Safety, and was, at the time of signing, the custodian of the records of the aforementioned laboratory and the signature was made in my official capacity as Scientific Director Designee of the State Breath Alcohol Testing Program.

(b) The verification of calibration for the above instrument has been received and reviewed. Instrument operator, Robert J Ely , whose name appears in my records, is qualified to verify the calibration of the above instrument.

(c) This instrument is certified for evidentiary use and these records are being retained as original documents in my custody.

Signed and sealed,

02/22/08

Jeanne Swartz — Date
Scientific Director Designee
State of Alaska
State Breath Alcohol Testing Program

STATE OF ALASKA SCIENTIFIC CRIME DETECTION LABORATORIES

BT9 03/00

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument
State of Alaska
Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130191

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130191

JANUARY 22, 2008 CPB 2/19/08

OPERATOR'S NAME:
ELY, RJ
OPERATOR'S NUMBER: 2681
SUBJECT'S LAST NAME:
VERIFICATION OF CAL
SUBJECT'S FIRST NAME/MI :
VER OF CAL
O.L. #: 123456789
DEPT/AGENCY: CR61
CASE/REPORT: 2008-XXXX
TEST TYPE: V
ALCO TARGET VALUE: .081
ALCO S/N: X172967

--- BREATH ANALYSIS ---

.081 ADJUSTED FOR 30.59 in

ALCO TARGET	.082	13:33
BLANK TEST	.000	13:34
INTERNAL STANDARD	VERIFIED	13:34
ALCO TV 30.59 in	.086	13:34
BLANK TEST	.000	13:35
SUBJECT SAMPLE	.000 CPB 2/19/08	13:35
BLANK TEST	.000	13:36
ALCO TV 30.59 in	.087	13:37
BLANK TEST	.000	13:37

20/55/50

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130191

JANUARY 22, 2008
TIME 13:28

--- DIAGNOSTIC CHECK ---

COMPUTER:	OKAY
PROGRAM:	OKAY
SOFTWARE DATE:	02/20/01
HEATERS	
SAMPLE CHAMBER:	49c
BREATH TUBE:	41c
BAROMETER:	30.59 in
FLOW DETECTOR:	OKAY
PUMP	
HIGH SPEED:	OKAY
DETECTOR:	OKAY
FILTERS:	OKAY
QUARTZ STANDARD:	OKAY
CALIBRATION:	OKAY

PRINTER TEST
!"#$%&'()*+,-./0123456789:;<=>?@ABCDEFG
HIJKLMNOPQRSTUVWXYZ[\]^_`abcdefghijklmno
pqrstuvwxyz{|}~