

# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Supervisor/Operator Performing the Verification Procedure:		DataMaster cdm S/N <u>130194</u>	
A	Name <u>Lucas P Amidon</u>	ID# <u>5651</u> Date <u>6/17/08</u> ✓	
	Agency <u>North Slope Borough PD</u> ✓	Phone # <u>907 852 6111</u>	
	Instrument Location <u>Barrow</u>		
B	Alco S/N <u>X 172978</u> ✓	Target Value <u>1.083</u> ✓ High Pressure <u>250 psi</u>	
	Alco Test Values	<u>1.084</u> ✓ 1 <sup>st</sup> Alco	<u>1.084</u> ✓ 2 <sup>nd</sup> Alco
	Signature <u>Lucas P Amidon</u>		<u>CSB</u> <u>6/20/08</u>
(OVER)			


Under the Alaska Rules of Evidence, I certify that (do not write in the area below)

(a) I am Jeanne Swartz, Forensic Scientist III, Scientific Crime Detection Laboratory, State of Alaska Department of Public Safety, and was, at the time of signing, the custodian of the records of the aforementioned laboratory and the signature was made in my official capacity as Scientific Director Designee of the State Breath Alcohol Testing Program.

(b) The verification of calibration for the above instrument has been received and reviewed. Instrument operator, Lucas Amidon, whose name appears in my records, is qualified to verify the calibration of the above instrument.

(c) This instrument is certified for evidentiary use and these records are being retained as original documents in my custody.

Signed and sealed,



Jeanne Swartz

Jeanne Swartz  
Scientific Director Designee  
State of Alaska  
State Breath Alcohol Testing Program

6/20/08

Date

JUN 10 2008

