

# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Supervisor/Operator Performing the Verification Procedure: <u>DataMaster cdm S/N 130194</u> ✓		
Name <u>JAMES MICHELS</u>	ID# <u>6502</u> Date <u>12-29-08</u> ✓	
A Agency <u>USBPD - BARROW</u>	Phone # <u>852-6111</u>	
Instrument Location <u>SQUAD ROOM - BARROW</u> ✓		
B Alco S/N <u>X172971</u> ✓	Target Value <u>.081</u> ✓ High Pressure <u>1100</u> <u>psi</u>	
Alco Test Values	<u>.084</u> ✓ 1 <sup>st</sup> Alco	<u>.085</u> ✓ 2 <sup>nd</sup> Alco
Signature <u>SGT James Michels SFS</u> ✓		<u>CSB</u> <u>1/6/09</u>
(OVER)		


Under the Alaska Rules of Evidence, I certify that, (Do Not write in the area below)

(a) I am Jeanne Swartz, Forensic Scientist III, Scientific Crime Detection Laboratory, State of Alaska Department of Public Safety, and was, at the time of signing, the custodian of the records of the aforementioned laboratory and the signature was made in my official capacity as Scientific Director Designee of the State Breath Alcohol Testing Program.

(b) The verification of calibration for the above instrument has been received and reviewed. Instrument operator, James Michels, whose name appears in my records, is qualified to verify the calibration of the above instrument.

(c) This instrument is certified for evidentiary use and these records are being retained as original documents in my custody.

Signed and sealed,



Jeanne Swartz  
Jeanne Swartz  
Scientific Director Designee  
State of Alaska  
State Breath Alcohol Testing Program

1/6/09  
Date

JAN 02 2009

# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130194 ✓

(CONTINUED FROM FRONT PAGE)

C

TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

## NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130194

SEPTEMBER 29, 2008 ✓

OPERATOR'S NAME: *CAS 10/09*

NIC-ELS/JAMES/D

OPERATOR'S NUMBER: 6532

SUBJECT'S LAST NAME:

VOC

SUBJECT'S FIRST NAME/MI :

VOC

O.L. #: VOC

DEPT/AGENCY: BRM1

CASE/REPORT: VOC

TEST TYPE: V

ALCO TARGET VALUE: 001 ✓

ALCO S/N: X172971

--- BREATH ANALYSIS ---

001 ADJUSTED FOR 30.71 in  
ALCO TARGET .003 22:06  
BLANK TEST .000 22:07  
INTERNAL STANDARD VERIFIED 22:08  
ALCO TV 30.71 in 004 ✓ 22:08  
BLANK TEST .000 22:09  
SUBJECT SAMPLE *CAS 10/09* .000 22:09  
BLANK TEST .000 22:10  
ALCO TV 30.72 in 005 ✓ 22:10  
BLANK TEST .000 22:11

## DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130194

SEPTEMBER 29, 2008 ✓  
TIME 21:51

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY

PROGRAM: OKAY

SOFTWARE DATE: 02/20/01

HEATERS

SAMPLE CHANGER: 49c

BREATH TUBE: 42c

BAROMETER: 30.71 in

FLOW DETECTOR: OKAY

PUMP

HIGH SPEED: OKAY

DETECTOR: OKAY

FILTERS: OKAY

QUARTZ STANDARD: OKAY

CALIBRATION: OKAY

PRINTER TEST

"\*0123456789:;<=>@ABCDEFGHIJKL  
MNPQRSTUVWXYZ[^\\_`abcde fghijklmnop  
qrstuvwxyz{|}~)"