

DEC 2 7 30

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

1/26/11 *11051* *130194* *EMB* *12/20/10*

Supervisor/Operator Performing the Verification Procedure: **DataMaster cdm S/N** 130291

Name Jaime Smith ID# 6762 Date 12/18/10

A Agency Denali National Park Phone # 907-683-9525

Instrument Location Denali National Park Headquarters

B Alco S/N X301591 Target Value 0.080 High Pressure 1250psi

Alco Test Values .080 .080
1st Alco 2nd Alco

Signature *Jaime Smith* *EMB* *12/30/10*

(OVER)

(Do Not write in the area below)


I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

Nita J. Bolz
Nita J. Bolz
Scientific Director
State Breath Alcohol Program

Subscribed and sworn before me this 9th day of MAR, 2010.

Carolyn M. Noland (Notary Seal Stamp)
Carolyn M. Noland DANIEL NICHOLAS
Notary Public, State of Alaska
Commission Expires with Office



VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

2009
11/23/10
11/20/10

DataMaster cdm S/N

130194
130204
0ms
12/30/10

(CONTINUED FROM FRONT PAGE)

C

TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130194

DECEMBER 18, 2010

OPERATOR'S NAME:

SMITH

OPERATOR'S NUMBER: 6752

ELEMENT'S LAST NAME:

VERIFICATION OF CAL

SUBJECT'S FIRST NAME-AT :

X

D.L. # : X

DEPT/AGENCY: DNLB

CASE/REPORT:

TEST TYPE: V

BLOOD TARGET VALUE: .030

BLOOD S/N: X301551

--- BREATH ANALYSIS ---

.000 ADJUSTED FOR 28.42 in

BLOOD TARGET .076 00:11

BLANK TEST .000 00:13

INTERNAL STANDARD VERIFIED 00:13

BLOOD TV 28.42 in .000 00:13

BLANK TEST .000 00:14

SUBJECT SAMPLE .000 00:14

BLANK TEST .000 00:15

BLOOD TV 28.42 in .000 00:15

BLANK TEST .000 00:16

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130194

DECEMBER 18, 2010

TIME 09:22

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY

PROGRAM: OKAY

SOFTWARE DATE: 01/08/09

HEATERS

SAMPLE CHAMBER: 43c

BREATH TUBE: 41c

BAROMETER: 28.43 in

FLOW DETECTOR: OKAY

PUMP

HIGH SPEED: OKAY

DETECTOR: OKAY

FILTERS: OKAY

QUARTZ STANDARD: OKAY

CALIBRATION: OKAY

PRINTER TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
SCIENTIFIC CRIME DETECTION LABORATORY
STATEWIDE BREATH ALCOHOL PROGRAM