

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130195

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130195

MARCH 22, 2011

CTF 3/31/11

OPERATOR'S NAME:

ELY, RJ

OPERATOR'S NUMBER:

SUBJECT'S NAME:

ALCOHOL TARGET VALUE:

ALCOHOL S/N: 75509

TEST DATE:

TEST TIME:

ALCOHOL TARGET VALUE: .079

ALCOHOL S/N: 75509

--- BREATH ANALYSIS ---

.079 ADJUSTED FOR 30.03 in
ALCO TARGET .079 11:45
BLANK TEST .000 11:46
INTERNAL STANDARD VERIFIED 11:46
ALCO TU 30.03 in .081 11:46
BLANK TEST .000 11:47
SUBJECT SAMPLE .030 11:47
BLANK TEST .000 11:48
ALCO TU 30.03 in .081 11:48
BLANK TEST .000 11:49

CTF
3/31/11

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130195

MARCH 22, 2011

TIME 11:43

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY

PROGRAM: OKAY

SOFTWARE DATE: 02/20/01

HEATERS

SAMPLE CHAMBER: 49c

BREATH TUBE: 42c

BAROMETER: 30.03 in

FLOW DETECTOR: OKAY

PUMP

HIGH SPEED: OKAY

DETECTOR: OKAY

FILTERS: OKAY

QUARTZ STANDARD: OKAY

CALIBRATION: OKAY

PASTURE TEST

ALCOHOL TARGET VALUE: .079

ALCOHOL S/N: 75509

TEST DATE: 03/22/11

TEST TIME: 11:43