

# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

JUN 10 2009

DataMaster cdm S/N <u>130195</u> ✓	
Supervisor/Operator Performing the Verification Procedure:	
A Name <u>Robert J. Ely</u>	ID# <u>2681</u> Date <u>06-09-2009</u> ✓
Agency <u>Craig Police Department</u>	Phone # <u>(907) 826-3330</u>
Instrument Location <u>506 2<sup>nd</sup> Street, Craig Alaska 99921</u>	✓
Alco S/N <u>77286</u> ✓	Target Value <u>.079</u> ✓ High Pressure <u>250</u> PSI
Alco Test Value Average <u>.083</u> ✓	<u>.083</u> ✓
1 <sup>st</sup> Alco	2 <sup>nd</sup> Alco
Signature <u>Robert J. Ely</u>	✓

(OVER)

(Do Not write in the area below)

BMB  
6/16/09

I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.


Nita J. Bolz

Nita J. Bolz  
Scientific Director  
State Breath Alcohol Program

Subscribed and sworn before me this 24<sup>th</sup> day of June 2009.

Carolyn M. Noland (Notary Seal Stamp)

Carolyn M. Noland  
Notary Public, State of Alaska  
Commission Expires with Office



# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130195 ✓

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

## NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130195  
JUNE 09, 2009  
OPERATOR'S NAME:  
ELY, RJ  
OPERATOR'S NUMBER: 2681  
SUBJECT'S LAST NAME:  
VERIFICATION OF CAL  
SUBJECT'S FIRST NAME/MI :  
VER OF CAL  
O.L. #: 123456789  
DEPT/AGENCY: CR61  
CASE/REPORT: 09-XXXX  
TEST TYPE: U  
ALCO TARGET VALUE: .079 ✓  
ALCO S/N: 77286 ✓

— BREATH ANALYSIS —

.079 ADJUSTED FOR 30.19 in ✓  
ALCO TARGET .079 12:41  
BLANK TEST .000 12:42  
INTERNAL STANDARD VERIFIED 12:42  
ALCO TO 30.19 in .003 12:42  
BLANK TEST .000 12:43  
SUBJECT SAMPLE .000 12:44  
BLANK TEST .000 12:44  
ALCO TO 30.19 in .003 ✓ 12:45  
BLANK TEST .000 12:45

*Handwritten:* 6/16/09

## DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130195  
JUNE 09, 2009 ✓  
TIME 12:38

— DIAGNOSTIC CHECK —

COMPUTER: OKAY  
PROGRAM: OKAY  
SOFTWARE DATE: 02/20/01  
HEATERS  
SAMPLE CHAMBER: 49c  
BREATH TUBE: 41c  
BAROMETER: 30.19 in  
FLOW DETECTOR: OKAY  
PUMP  
HIGH SPEED: OKAY  
DETECTOR: OKAY  
FILTERS: OKAY  
QUARTZ STANDARD: OKAY  
CALIBRATION: OKAY

PRINTER TEST  
!"#\$%&'()\*+,-./0123456789:;<=>?@ABCDEF  
HIJKLMNPOQRSTUVWXYZ[\]^\_`abcdefghijklmnop  
qrstuvwxyz{|}~@