

**VERIFICATION OF CALIBRATION REPORT** JUN 22 2009

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130195 ✓

Supervisor/Operator Performing the Verification Procedure:

A Name Robert J. Ely ID# 2681 Date 06-18-2009 ✓

Agency Craig Police Department Phone # (907) 826-3330

Instrument Location 506 2<sup>nd</sup> Street, Craig Alaska 99921 ✓

---

Alco S/N 77286 ✓ Target Value .079 ✓ High Pressure 200 PSI

Alco Test Value Average .083 ✓

1<sup>st</sup> Alco

.083 ✓

2<sup>nd</sup> Alco

Signature *Robert J. Ely* ✓

(OVER)

*BMB*  
*6/25/09*  
*89*  
*6/25/09*

(Do Not write in the area below)

I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.


*Nita J. Bolz*

Nita J. Bolz  
Scientific Director  
State Breath Alcohol Program

Subscribed and sworn before me this 20<sup>th</sup> day of June, 2009.

*Carolyn M. Noland* (Notary Seal Stamp)

Carolyn M. Noland  
Notary Public, State of Alaska  
Commission Expires with Office



# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130195

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

## NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130195

JUNE 18, 2009

OPERATOR'S NAME:  
ELY, RJ

OPERATOR'S NUMBER: 2681

SUBJECT'S LAST NAME:  
VERIFICATION OF CAL

SUBJECT'S FIRST NAME/NI :  
VERIF CAL

O.L. #: 123456789

DEPT/AGENCY: CR81

CASE/REPORT: 2009-XXXX

TEST TYPE: U

ALCO TARGET VALUE: .079 ✓

ALCO S/N: 77286

— BREATH ANALYSIS —

.079 ADJUSTED FOR 30.14 in		
ALCO TARGET	.079	11:57
BLANK TEST	.000	11:58
INTERNAL STANDARD	VERIFIED	11:58
ALCO TV 30.14 in	.063	11:58
BLANK TEST	.000	11:59
SUBJECT SAMPLE	.000	12:00
BLANK TEST	.000	12:00
ALCO TV 30.14 in	.063	12:01
BLANK TEST	.000	12:02

*Handwritten:* DMG 6/25/09

## DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130195

JUNE 18, 2009 ✓  
TIME 11:55

— DIAGNOSTIC CHECK —

COMPUTER: OKAY

PROGRAM: OKAY

SOFTWARE DATE: 02/20/01

HEATERS

SAMPLE CHAMBER: 4c

BREATH TUBE: 41c

BAROMETER: 30.14 in

FLOW DETECTOR: OKAY

PUMP

HIGH SPEED: OKAY

DETECTOR: OKAY

FILTERS: OKAY

QUARTZ STANDARD: OKAY

CALIBRATION: OKAY

PRINTER TEST

!@#\$%^&\*~.-/0123456789:;<=>?@BCDEFG  
HIJKLMNOPQRSTUVWXYZ\|\_`abcde fghijklmnop  
qrstuvwxyz{|}~@