

# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

FEB 12 2010

|  |   |
|--|---|
| DataMaster cdm S/N <u>130195</u>   |   |
| Supervisor/Operator Performing the Verification Procedure:               |   |
| Name <u>Robert J. Ely</u>  | ID# <u>2681</u> Date <u>02-10-2010</u>                |
| Agency <u>Craig Police Department</u>                                    | Phone # <u>(907) 826-3330</u>                         |
| Instrument Location <u>506 2<sup>nd</sup> Street, Craig Alaska 99921</u> |   |
| Alco S/N <u>77286</u>  | Target Value <u>.079</u> High Pressure <u>900</u> PSI |
| Alco Test Value Average <u>.083</u>                                      | <u>.082</u>   |
| 1 <sup>st</sup> Alco   | 2 <sup>nd</sup> Alco                                  |
| Signature <u>Robert J. Ely</u>   |   |

(OVER)

(Do Not write in the area below)

*RMB*  
*2/23/2010*


I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

Nita J. Bolz  
Nita J. Bolz  
Scientific Director  
State Breath Alcohol Program

Subscribed and sworn before me this 24<sup>th</sup> day of February, 2010.

Carolyn M. Noland  
(Notary Seal Stamp)  
Carolyn M. Noland  
Notary Public, State of Alaska  
Commission Expires with Office



# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130195

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

## NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130195

FEBRUARY 10, 2010

OPERATOR'S NAME: *ELV, RJ*  
ELV, RJ  
OPERATOR'S NUMBER: 2601  
SUBJECT'S LAST NAME:  
VERIFICATION OF CAL  
SUBJECT'S FIRST NAME/MI :  
VER OF CAL  
O.L. #: 123456789  
DEPT/AGENCY: CR61  
CASE/REPORT: 2010-XXXX  
TEST TYPE: V  
ALCO TARGET VALUE: .079 ✓  
ALCO S/N: 77286 ✓

### BREATH ANALYSIS

.079 ADJUSTED FOR 29.78 in  
ALCO TARGET .078 12:17  
BLANK TEST .000 12:18  
INTERNAL STANDARD VERIFIED 12:18  
ALCO TV 29.78 in .083 ✓ 12:19  
BLANK TEST .000 12:19  
SUBJECT SAMPLE .000 12:20  
BLANK TEST .000 12:21  
ALCO TV 29.78 in .082 ✓ 12:21  
BLANK TEST .000 12:22

## DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130195

FEBRUARY 10, 2010  
TIME 12:15

### DIAGNOSTIC CHECK

COMPUTER: OKAY  
PROGRAM: OKAY  
SOFTWARE DATE: 02/20/01  
HEATERS  
SAMPLE CHAMBER: 49c  
BREATH TUBE: 40c  
BAROMETER: 29.78 in  
FLOW DETECTOR: OKAY  
PUMP  
HIGH SPEED: OKAY  
DETECTOR: OKAY  
FILTERS: OKAY  
QUARTZ STANDARD: OKAY  
CALIBRATION: OKAY

### PRINTER TEST

!"#\$%&'()\*+,-./0123456789:;<=?@ABCDEFGHIJKL  
MNPQRSTUVWXYZ[\]^\_`abcdefghijklmnopqrstu  
vwxyz{|}~■