

VERIFICATION OF CALIBRATION REPORT

*of DataMaster cdm Breath Test Instrument
State of Alaska*

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N		130195 ✓
Supervisor/Operator Performing the Verification Procedure:		
A Name	Robert J. Ely	ID# 2681 ✓ Date 09-01-2010 ✓
Agency	Craig Police Department	Phone # (907) 826-3330
Instrument Location 506 2 nd Street, Craig Alaska 99921		
Alco S/N	X172978 ✓	Target Value .079 ✓ High Pressure 0 PSI
Alco Test Value Average	.083 ✓ 1 st Alco	.083 ✓ 2 nd Alco
Signature	<i>Robert J. Ely</i>	COB 9/7/10

(OVER)

(Do Not write in the area below)


I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

Nita J. Bolz
 Nita J. Bolz
 Scientific Director
 State Breath Alcohol Program

Subscribed and sworn before me this 8th day of DEC, 2010.

Daniel L. Nicholas (Notary Seal Stamp)
 Daniel L. Nicholas
 Notary Public, State of Alaska
 Commission Expires with Office



VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130195

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130195

SEPTEMBER 01, 2010

OPERATOR'S NAME: *COB 9/7/10*
ELY, RJ
OPERATOR'S NUMBER: 2681
SUBJECT'S LAST NAME:
VERIFICATION OF CAL
SUBJECT'S FIRST NAME/MI :
VER OF CAL
O.L. #: 123456789
DEPT/AGENCY: CR61
CASE/REPORT: 10-XXXXX
TEST TYPE: U
ALCO TARGET VALUE: .079
ALCO S/N: X172978

— BREATH ANALYSIS —

.079 ADJUSTED FOR 30.17 in
ALCO TARGET .079 14:11
BLANK TEST .000 14:12
INTERNAL STANDARD VERIFIED 14:12
ALCO TV 30.17 in .063 14:12
BLANK TEST .000 14:13
SUBJECT SAMPLE *COB 9/7/10* .000 14:14
BLANK TEST .000 14:15
ALCO TV 30.17 in .063 14:15
BLANK TEST .000 14:16

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130195

SEPTEMBER 01, 2010
TIME 14:07

— DIAGNOSTIC CHECK —

COMPUTER: OKAY
PROGRAM: OKAY
SOFTWARE DATE: 02/20/01
HEATERS
SAMPLE CHAMBER: 49c
BREATH TUBE: 42c
BAROMETER: 30.17 in
FLOW DETECTOR: OKAY
PUMP
HIGH SPEED: OKAY
DETECTOR: OKAY
FILTERS: OKAY
QUARTZ STANDARD: OKAY
CALIBRATION: OKAY

PRINTER TEST

!"#\$%&'()*+,-./0123456789:;<=>?@ABCDEF6
HIJKLMNOPQRSTUVWXYZ\]^_`abcdefgijklmnop
qrstuvwxyz{|}~■