

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130195

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130195

NOVEMBER 13, 2010

OPERATOR'S NAME:

ELY, RJ

OPERATOR'S NUMBER: 2581

SUBJECT'S LAST NAME:

VERIFICATION OF CAL

SUBJECT'S FIRST NAME/MI :

IER OF CAL

D.L. #: 123456789

DEPT/AGENCY: 2201

CASE REPORT: 10-XXXX

TEST TYPE: U

ALCO TARGET VALUE: .078

ALCO S/N: X173355

--- BREATH ANALYSIS ---

.078 ADJUSTED FOR 30.32 in
ALCO TARGET .079 09:51
BLANK TEST .000 09:51
INTERNAL STANDARD VERIFIED 09:52
ALCO TV 30.32 in .000 09:52
BLANK TEST .000 09:53
SUBJECT SAMPLE .200 09:53
BLANK TEST .000 09:54
ALCO TV 30.32 in .000 09:54
BLANK TEST .000 09:55

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130195

NOVEMBER 13, 2010

TIME 09:49

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY

PROGRAM: OKAY

SOFTWARE DATE: 02/28/01

HEATERS

SAMPLE CHAMBER: 49c

BREATH TUBE: 41c

BAROMETER: 30.32 in

FLOW DETECTOR: OKAY

PUMP

HIGH SPEED: OKAY

DETECTOR: OKAY

FILTERS: OKAY

QUARTZ STANDARD: OKAY

CALIBRATION: OKAY

PRINTER TEST

!@#%&'()*+,-./0123456789:;<=>?@ABCDEF
GHIJKL/MNOPQRSTUVWXYZ{|}~' abcdefghijklmnop
qrstuvwxyz{|}~@