

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

DEC 29 2010

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Supervisor/Operator Performing the Verification Procedure: DataMaster cdm S/N 130286

Name D.C. Stan Swetkof ID# 5002 Date 12/26/10

A Agency North Pole PD Phone # 907-488-0612

Instrument Location 125 Snowman Lane, North Pole, AK 99705

B Alco S/N 124662 Target Value .087 High Pressure 600 PSI

Alco Test Values .084 .083
1st Alco 2nd Alco

Signature [Signature] [Signature] [Signature]

(OVER)

(Do Not write in the area below)


I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

Nita J. Bolz
Scientific Director
State Breath Alcohol Program

Subscribed and sworn before me this 23rd day of FEB, 2010
(Notary Seal Stamp)

Daniel L. Nicholas
Notary Public, State of Alaska
Commission Expires with Office



VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130286

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130286

DECEMBER 26, 2010

OPERATOR'S NAME:
SMETZOF/STANLEY
OPERATOR'S NUMBER: 5002
SUBJECT'S LAST NAME:
VERIFICATION OF CAL
SUBJECT'S FIRST NAME-AL:

SEC. #: L
DEPT/AGENCY: HP-1
CASE/REPORT:
TEST TYPE: U
ALCO TARGET VALUE: .087
ALCO S/N: X124652

--- BREATH ANALYSIS ---

.087 ADJUSTED FOR 29.39 in
ALCO TARGET .085 00:51
BLANK TEST .000 00:51
INTERNAL STANDARD VERIFIED 00:51
ALCO TO 29.39 in .084 00:52
BLANK TEST .000 00:53
SUBJECT SAMPLE .080 00:53
BLANK TEST .000 00:54
ALCO TO 29.39 in .083 00:55
BLANK TEST .000 00:55

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130286

DECEMBER 26, 2010
TIME 08:47

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY
PROGRAM: OKAY
SOFTWARE DATE: 01/08/09
HEATERS
SAMPLE CHAMBER: 49c
BREATH TUBE: 41c
BAROMETER: 29.38 in
FLOW DETECTOR: OKAY
PUMP
HIGH SPEED: OKAY
DETECTOR: OKAY
FILTERS: OKAY
QUARTZ STANDARD: OKAY
CALIBRATION: OKAY

PRINTER TEST

!@#%&'()*+,-./0123456789:;<=>?@BCDEFG
HIJKL;NOPQRSTUUVWXYZ[]^_`abcdefgijklmnop
qrstuvwxyz{|}~