MAY 0 4 2009

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument
State of Alaska
Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130287

Supervisor/Operator Performing the Verification Procedure:

A

Name Kelly George ID# 4447 Date 5-4-09

Agency Kenai Police Department Phone # 283-7879

Instrument Location Wildwood Pretrial Facility

B

Alco S/N X 301542 Target Value .082 High Pressure 1000

Alco Test Values .082 (1st Alco) .081 (2nd Alco)

Signature Kelly George

(OVER)

(Do Not write in the area below)

I, **Nita J. Bolz**, after being first duly sworn, depose and state as follows:

(1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
(2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
(3) I am the Scientific Director of the State Breath Alcohol Program.
(4) In that capacity, I am **responsible for overseeing the breath alcohol program, including assuring that persons responsible** for verifying **the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.**
(5) The attached verification is [redacted] regularly [redacted] **verification.**

Nita J. Bolz
Scientific Director
State Breath Alcohol Program

Subscribed and sworn before me this 22 day of May 2009.

(Notary Seal Stamp)

Carolyn M. Noland
Notary Public, State of Alaska
Commission Expires with Office

Carolyn M. Noland NOTARY PUBLIC STATE OF ALASKA State of Alaska

BT9 03/00

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130287

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST



DIAGNOSTIC CHECK