

# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument  
State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N <u>130287</u>		
Supervisor/Operator Performing the Verification Procedure:		
Name <u>Kelly George</u>	ID# <u>4447</u> Date <u>3-20-11</u>	
A Agency <u>Kenai Police Dept.</u>	Phone # <u>283-7879</u>	
Instrument Location <u>Wildwood Pretrial Jail</u>		
Alco S/N <u>+301600</u> Target Value <u>.082</u> High Pressure <u>700</u>		
B Alco Test Values	<u>.079</u> 1st Alco	<u>.080</u> 2 <sup>nd</sup> Alco
Signature <u>Kelly George</u>		
(OVER)		

DMB  
5/25/11

(Do Not write in the area below)

I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

Nita J. Bolz  
Nita J. Bolz  
Scientific Director  
State Breath Alcohol Program

Subscribed and sworn before me this 27 day of MAY, 2010.

NJB  
2011

Daniel L. Nicholas (Notary Seal Stamp)  
Daniel L. Nicholas  
Notary Public, State of Alaska  
Commission Expires with Office



# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130287

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

## NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130287

MARCH 20, 2011

OPERATOR'S NAME:

GEORGE/KELLY

OPERATOR'S NUMBER: 4447

SUBJECT'S LAST NAME:

UOC

SUBJECT'S FIRST NAME/NI:

UOC

O.L. #: UOC

DEPT/AGENCY: KENI

CASE/REPORT: UOC

TEST TYPE: U

ALCO TARGET VALUE: .082

ALCO S/N: K301600

### --- BREATH ANALYSIS ---

.082 ADJUSTED FOR 30.09 in

ALCO TARGET	.082	01:37
BLANK TEST	.000	01:38
INTERNAL STANDARD	VERIFIED	01:38
ALCO TU 30.09 in	.079	01:39
BLANK TEST	.000	01:39
SUBJECT SAMPLE	.000	01:41
BLANK TEST	.000	01:42
ALCO TU 30.09 in	.000	01:42
BLANK TEST	.000	01:43

## DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130287

MARCH 20, 2011

TIME 01:35

### --- DIAGNOSTIC CHECK ---

COMPUTER:	OKAY
PROGRAM:	OKAY
SOFTWARE DATE:	02/20/01
HEATERS	
SAMPLE CHAMBER:	48c
BREATH TUBE:	42c
BAROMETER:	30.09 in
FLOW DETECTOR:	OKAY
PUMP	
HIGH SPEED:	OKAY
DETECTOR:	OKAY
FILTERS:	OKAY
QUARTZ STANDARD:	OKAY
CALIBRATION:	OKAY

### PRINTER TEST

!"#\$%&'()\*+,-./0123456789:;<=>?@ABCDEF  
GHIJKLMNOPQRSTUVWXYZ[\]^\_`abcdefgijklmno  
pqrstuvwxyz{|}~■