

MAR 04 2009

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Supervisor/Operator Performing the Verification Procedure: DataMaster cdm S/N 130288 -

Name DIXIE SPENCER ID# 2280 Date 3/2/09

A Agency AST ANIAK Phone # 675-4398

Instrument Location ANIAK

B Alco S/N 56799 Target Value .084 High Pressure 1150

Alco Test Values .085 ✓ .085 ✓
 1st Alco 2nd Alco

Signature Dixie Spencer ✓

003
3/24/09

(OVER)

Under the Alaska Rules of Evidence, I certify that, ^(Do Not write in the area below)

(a) I am Jeanne Swartz, Forensic Scientist III, Scientific Crime Detection Laboratory, State of Alaska Department of Public Safety, and was, at the time of signing, the custodian of the records of the aforementioned laboratory and the signature was made in my official capacity as Scientific Director Designee of the State Breath Alcohol Testing Program.

(b) The verification of calibration for the above instrument has been received and reviewed. Instrument operator, Dixie Spencer, whose name appears in my records, is qualified to verify the calibration of the above instrument.

(c) This instrument is certified for evidentiary use and these records are being retained as original documents in my custody.

Signed and sealed,



Jeanne Swartz

Jeanne Swartz Date 3/24/09
Scientific Director Designee
State of Alaska
State Breath Alcohol Testing Program

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130288

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130288

MARCH 02, 2009

OPERATOR'S NAME:
SPENCER DIXIE W
OPERATOR'S NUMBER: 2200
SUBJECT'S LAST NAME:
HHHH

SUBJECT'S FIRST NAME/NI :
JJJJ

D.L. #: 1234567
DEPT/AGENCY: AN10
CASE/REPORT: 09-
TEST TYPE: U
ALCO TARGET VALUE: .004
ALCO S/N: 56799

--- BREATH ANALYSIS ---

.004 ADJUSTED FOR 29.85 in
ALCO TARGET .003 12:52
BLANK TEST .000 12:53
INTERNAL STANDARD VERIFIED 12:53
ALCO TU 29.85 in .005 12:54
BLANK TEST .000 12:55
SUBJECT SAMPLE .000 12:56
BLANK TEST .000 12:57
ALCO TU 29.85 in .005 12:57
BLANK TEST .000 12:58

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130288

MARCH 02, 2009
TIME 12:59

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY
PROGRAM: OKAY
SOFTWARE DATE: 02/20/01
HEATERS
SAMPLE CHAMBER: 49c
BREATH TUBE: 41c
BAROMETER: 29.85 in
FLAM DETECTOR: OKAY
PUMP
HIGH SPEED: OKAY
DETECTOR: OKAY
FILTERS: OKAY
QUARTZ STANDARD: OKAY
CALIBRATION: OKAY

PRINTER TEST

!@#%&'()*+,-./0123456789:;<=>?@ABCDEFGHIJ
KLMNOPQRSTUVWXYZ[\]^_`abcde fghijklmnop
qrstuvwxyz{|}~