VERIFICATION OF CALIBRATION REPORT

of Datamaster cdm
State of Alaska
Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Datamaster cdm S/N 130290

Supervisor/Operator Performing the Verification Procedure:

A

Name C. RITALA ID# 3791 Date 7/22/08

Agency Anchorage P.D. Phone # 786-2640

Instrument Location Wollam Sub (Eagle River)

B

Alco S/N 58427 Target Value 082 High Pressure 1150

Alco Test Value Average 085 (1st Alco) 083 (2nd Alco)

Signature [signature] / 1349 7/28/08

(OVER)

(Do Not write in the area below)

Under the Alaska Rules of Evidence, I certify that;

(a) I am Jeanne Swartz, Forensic Scientist III, Scientific Crime Detection Laboratory, State of Alaska Department of Public Safety, and was, at the time of signing, the custodian of the records of the aforementioned laboratory and the signature was made in my official capacity as Scientific Director Designee of the State Breath Alcohol Testing Program.

(b) The verification of calibration for the above instrument has been received and reviewed. Instrument operator, Chris Ritala, whose name appears in my records, is qualified to verify the calibration of the above instrument.

(c) This instrument is certified for evidentiary use and these records are being retained as original documents in my custody.

Signed and sealed,

STATE OF ALASKA SCIENTIFIC CRIME DETECTION LABORATORY

[signature] 7/30/08

Jeanne Swartz Date
Scientific Director Designee
State of Alaska
State Breath Alcohol Testing Program

BT9 01/00

VERIFICATION OF CALIBRATION REPORT

of Datamaster cdm
State of Alaska
Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Datamaster cdm S/N 130290

(CONTINUED FROM FRONT PAGE)

C. TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130290

JULY 22, 2008 CDS 7/28/08

OPERATOR'S NAME:
RITALA
OPERATOR'S NUMBER: 3791
SUBJECT'S LAST NAME:
VERIFICATION OF CALIBRATION
SUBJECT'S FIRST NAME/MI :
A
O.L. #: A
DEPT/AGENCY: ANC1
CASE/REPORT: A
TEST TYPE: V
ALCO TARGET VALUE: .082
ALCO S/N: 58427

— BREATH ANALYSIS —

.082: ADJUSTED FOR 29.75 in

ALCO TARGET	.081	19:02
BLANK TEST	.000	19:03
INTERNAL STANDARD	VERIFIED	19:03
ALCO TV 29.75 in	.085	19:04
BLANK TEST	.000	19:05
SUBJECT SAMPLE	.000	19:05
BLANK TEST	.000	19:06
ALCO TV 29.75 in	.083	19:07
BLANK TEST	.000	19:07

CDS 7/28/08

CR/1349 7/22/08

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130290

JULY 22, 2008
TIME 18:59

— DIAGNOSTIC CHECK —

COMPUTER:	OKAY
PROGRAM:	OKAY
SOFTWARE DATE:	02/20/01
HEATERS	
SAMPLE CHAMBER:	49c
BREATH TUBE:	[illegible]
BAROMETER:	29.73 in
FLOW DETECTOR:	OKAY
PUMP	
HIGH SPEED:	OKAY
DETECTOR:	OKAY
FILTERS:	OKAY
QUARTZ STANDARD:	OKAY
CALIBRATION:	OKAY

PRINTER TEST
!"#$%&'()*+,-./0123456789:;<=>?@ABCDEFG
HIJKLMNOPQRSTUVWXYZ[\]^_`abcdefghijklmno
pqrstuvwxyz{|}~

CR/1349 7/22/08