

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

DEC 29 2009


Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Supervisor/Operator Performing the Verification Procedure: DataMaster cdm S/N <u>130290</u>	
A	Name <u>STEVE DUNN</u> ID# <u>4106</u> Date <u>12-23-09</u>
	Agency <u>APP</u> Phone # <u>764-1361</u>
	Instrument Location <u>ANCHORAGE JAIL</u>
B	Alco S/N <u>X301536</u> Target Value <u>.086</u> High Pressure <u>750 PSE.</u>
	Alco Test Value Average <u>.085</u> <u>.085</u>
	1 st Alco 2 nd Alco
Signature 	
083 1/13/10	
(OVER)	


(Do Not write in the area below)

I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.


Nita J. Bolz
Scientific Director
State Breath Alcohol Program

Subscribed and sworn before me this 26th day of January, 2010.


(Notary Seal Stamp)
Carolyn M. Noland
Notary Public, State of Alaska
Commission Expires with Office



VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130290

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

S. Dunn #1361

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130290

DECEMBER 23, 2009

OPERATOR'S NAME:

DUNN/S

OPERATOR'S NUMBER: 4106

SUBJECT'S LAST NAME:

VERIFICATION OF CAL

SUBJECT'S FIRST NAME/MI :

V

O.L. #: V

DEPT/AGENCY: ANCI

CASE/REPORT: V

TEST TYPE: V

ALCO TARGET VALUE: .056

ALCO S/N: X301536

--- BREATH ANALYSIS ---

.006 ADJUSTED FOR 29.85 in
ALCO TARGET .005 18:03
BLANK TEST .000 18:04
INTERNAL STANDARD VERIFIED 18:24
ALCO TV 29.85 in .005 ✓ 18:24
BLANK TEST .000 18:25
SUBJECT SAMPLE .000 18:25
BLANK TEST .000 18:26
ALCO TV 29.83 in .005 ✓ 18:27
BLANK TEST .000 18:27

COS 1/12/10

DIAGNOSTIC CHECK

S. Dunn #1361

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130290

DECEMBER 23, 2009

TIME 18:20

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY
PROGRAM: OKAY
SOFTWARE DATE: 01/08/09
HEATERS
SAMPLE CHAMBER: 49c
BREATH TUBE: 42c
BAROMETER: 29.83 in
FLOW DETECTOR: OKAY
PUMP
HIGH SPEED: OKAY
DETECTOR: OKAY
FILTERS: OKAY
QUARTZ STANDARD: OKAY
CALIBRATION: OKAY

PRINTER TEST

!@#%&'()*+,-./0123456789:;<=>?@ABCDEFGHI
HIJKL MNOPQRSTUVWXYZ[]^_`abcde fghijklmnop
qrstuvwxyz{|}~