

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

MAY 11 2009

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Supervisor/Operator Performing the Verification Procedure:	
DataMaster cdm S/N <u>130291</u> ✓	
Name <u>Sean Fielding</u>	ID# <u>5189</u> Date <u>4/30/09</u> ✓
A Agency <u>Denali National Park</u>	Phone # <u>683 9528</u>
Instrument Location <u>Denali NP - HQ</u> ✓	
Alco S/N <u>Y301553</u> ✓ Target Value <u>.083</u> ✓ High Pressure <u>700</u>	
B Alco Test Value Average <u>.086</u> ✓ <u>.086</u> ✓	
1 st Alco 2 nd Alco	
Signature <u>[Signature]</u> ✓	
CMB 3/13/09	
(OVER)	


I, Nita J. Bolz, after being first duly sworn, depose and state as follows: (Do not write in the area below)

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.

[Signature]
Nita J. Bolz
Scientific Director
State Breath Alcohol Program

Subscribed and sworn before me this 15 day of May, 2009.

[Signature] (Notary Seal Stamp)
Carolyn M. Noland
Notary Public, State of Alaska
Commission Expires with Office



VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130291

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
SCIENTIFIC CRIME DETECTION LABORATORY

APRIL 20, 2009 ✓ *CB 6/18/09*

OPERATOR'S NAME:
EYEGLASS BEARER
INSTRUMENT NUMBER: 5189
SUBJECT'S LAST NAME:
VERIFICATION OF CAL
SUBJECT'S FIRST NAME/INI:
H
O.L. #: 1274567
DEPT AGENCY: DNLB
CASE REPORT: 047009
TEST TYPE: U
ALCOHOL TEST VALUE: 0.00 ✓
TEST STRIP: 118

--- BREATH ANALYSIS ---

0.00 REQUESTED FOR 20.29 in ✓
ALCOHOL TEST 0.00 13:40
BLANK TEST 0.00 13:41
INTERNAL STANDARD (VERIFIED) 0.00 13:41
ALCOHOL TEST 0.00 ✓ 13:42
BLANK TEST 0.00 13:42
SUBJECT SAMPLE *CB 6/18/09* 0.00 13:43
BLANK TEST 0.00 13:44
ALCOHOL TEST 0.00 ✓ 13:44
BLANK TEST 0.00 13:45

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
SCIENTIFIC CRIME DETECTION LABORATORY

APRIL 20, 2009 ✓
TIME 13:40

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY
PROGRAM: OKAY
SOFTWARE DATE: 02/28/01
HEATERS
SAMPLE CHARGER: 50c
TREAT: 118
PROPAGATOR: 13:43
FLAM DETECTOR: OKAY
PUMP
HIGH SPEED: OKAY
DETECTOR: OKAY
FILTERS: OKAY
QUARTZ STANDARD: OKAY
CALIBRATION: OKAY

DRIVER TEST
INSTRUMENT NUMBER: 5189
HIGH SPEED TEST: 13:45
DATE: 4/20/09