

# VERIFICATION OF CALIBRATION REPORT

JUL 27 2010

of DataMaster cdm Breath Test Instrument  
State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Supervisor/Operator Performing the Verification Procedure:	
Name	<u>D. R. Sturkie</u> ID# <u>3559</u> Date <u>07-27-10</u> ✓
A Agency	<u>APD</u> Phone # <u>907-2672</u>
Instrument Location <u>Eagle River Substation</u>	
DataMaster cdm S/N <u>130292</u> ✓	
B Alco S/N	<u>75260</u> ✓ Target Value <u>.081</u> ✓ High Pressure <u>1100</u>
Alco Test Values <u>.081</u> ✓ 1st Alco <u>.081</u> ✓ 2 <sup>nd</sup> Alco	
Signature	<u>D. R. Sturkie</u> <span style="float: right;"><u>DB</u> 7/30/10</span>

(OVER)

(Do Not write in the area below)


I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

Nita J. Bolz  
Scientific Director  
State Breath Alcohol Program

Subscribed and sworn before me this 11<sup>th</sup> day of AUGUST, 2010.

Daniel L. Nicholas (Notary Seal Stamp)  
Daniel L. Nicholas  
Notary Public, State of Alaska  
Commission Expires with Office



# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument  
State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130292

JULY 27, 2010

OPERATOR'S NAME:

STU

OPERATOR'S NUMBER: <sup>3559</sup> ~~R~~ 7/27/10 *ow*

SUBJECT'S LAST NAME:

STURKIE

SUBJECT'S FIRST NAME/MI :

VOC

O.L. #: VOC

DEPT/AGENCY: ANC1

CASE/REPORT: VOC

TEST TYPE: V

ALCO TARGET VALUE: .081

ALCO S/N: 75260

--- BREATH ANALYSIS ---

.081 ADJUSTED FOR 30.01 in  
ALCO TARGET .081 05:17  
BLANK TEST .000 05:18  
INTERNAL STANDARD VERIFIED 05:18  
ALCO TV 30.01 in .081 05:19  
BLANK TEST .000 05:19  
SUBJECT SAMPLE <sup>CUB</sup> 7/30/10 .000 05:20  
BLANK TEST .000 05:21  
ALCO TV 30.01 in .081 05:21  
BLANK TEST .000 05:22

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130292

JULY 27, 2010

TIME 05:22

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY  
PROGRAM: OKAY  
SOFTWARE DATE: 01/08/09  
HEATERS  
SAMPLE CHAMBER: 48c  
BREATH TUBE: 43c  
BAROMETER: 30.01 in  
FLOW DETECTOR: OKAY  
PUMP  
HIGH SPEED: OKAY  
DETECTOR: OKAY  
FILTERS: OKAY  
QUARTZ STANDARD: OKAY  
CALIBRATION: OKAY

PRINTER TEST

!"#\$%&'()\*+,-./0123456789:;<=>?@ABCDEF  
HIJKLMNOPQRSTUVWXYZ[\]^\_`abcdefghijklmnopqr  
stuvwxyz{|}~△