VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument
State of Alaska
Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130293

Supervisor/Operator Performing the Verification Procedure:

A

Name Brian Mitchell ID# 4119 Date 7-2-08

Agency Alaska State Troopers Phone # 486-4121

Instrument Location Kodiak Post, Interview Room

B

Alco S/N X124626 Target Value .082 High Pressure 1100

Alco Test Value Average .083 (1st Alco) .086 (2nd Alco)

Signature [signature] CDB 7/10/08

(OVER)

Under the Alaska Rules of Evidence, I certify that;

(Do Not write in the area below)

(a) I am Jeanne Swartz , Forensic Scientist III, Scientific Crime Detection Laboratory, State of Alaska Department of Public Safety, and was, at the time of signing, the custodian of the records of the aforementioned laboratory and the signature was made in my official capacity as Scientific Director Designee of the State Breath Alcohol Testing Program.

(b) The verification of calibration for the above instrument has been received and reviewed. Instrument operator, Brian Mitchell , whose name appears in my records, is qualified to verify the calibration of the above instrument.

(c) This instrument is certified for evidentiary use and these records are being retained as original documents in my custody.

Signed and sealed,

[signature] 7/13/08

Jeanne Swartz Date
Scientific Director Designee
State of Alaska
State Breath Alcohol Testing Program

STATE OF ALASKA SCIENTIFIC CRIME DETECTION LABORATORY

BT9 03/00

JUL 07 2008

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument
State of Alaska
Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130293

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130293

JULY 02, 2008 CAB 7/10/08

OPERATOR'S NAME:
SUPERVISOR
OPERATOR'S NUMBER: 4119
SUBJECT'S LAST NAME:
VERIFICATION
SUBJECT'S FIRST NAME/MI :
VERIFICATION
D.L. #: X
DEPT/AGENCY: KOD0
CASE/REPORT: X
TEST TYPE: V
ALCO TARGET VALUE: .082
ALCO S/N: X124626

--- BREATH ANALYSIS ---

.082 ADJUSTED FOR 29.85 in

ALCO TARGET	.081	18:56
BLANK TEST	.000	18:57
INTERNAL STANDARD	VERIFIED	18:58
ALCO TV 29.85 in	.083	18:58
BLANK TEST	.000	18:59
SUBJECT SAMPLE	.000	18:59
BLANK TEST	.000	19:00
ALCO TV 29.85 in	.086	19:00
BLANK TEST	.000	19:01

CAB 7/10/08

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130293

JULY 02, 2008
TIME 19:02

--- DIAGNOSTIC CHECK ---

COMPUTER:	OKAY
PROGRAM:	OKAY
SOFTWARE DATE:	02/20/01
HEATERS	
SAMPLE CHAMBER:	49c
BREATH TUBE:	41c
BAROMETER:	29.85 in
FLOW DETECTOR:	OKAY
PUMP	
HIGH SPEED:	OKAY
DETECTOR:	OKAY
FILTERS:	OKAY
QUARTZ STANDARD:	OKAY
CALIBRATION:	OKAY

PRINTER TEST
!"#$%&'()*+,-./0123456789:;<=>?@ABCDEFG
HIJKLMNOPQRSTUVWXYZ[\]^_`abcdefghijklmno
pqrstuvwxyz{|}~