

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Supervisor/Operator Performing the Verification Procedure:		DataMaster cdm S/N	<u>130294</u>
Name		<u>David No11</u>	ID# <u>6590</u> Date <u>6/30/09</u>
A	Agency	<u>APD</u>	Phone # <u>786-8688</u>
Instrument Location <u>Eagle River Sub</u>			
Alco S/N		<u>X172946</u>	Target Value <u>0.084</u> High Pressure <u>1100</u>
Alco Test Value Average		<u>0.082</u>	<u>0.082</u>
		1 st Alco	2 nd Alco
Signature <u>[Signature]</u>			<u>008</u> <u>8/6/09</u>

(OVER)

I, Nita J. Bolz, after being first duly sworn, depose and state as follows (State as follows in the area below)

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

[Signature]
Nita J. Bolz
Scientific Director
State Breath Alcohol Program

Subscribed and sworn before me this 12th day of Aug, 2009.

[Signature] (Notary Seal Stamp)
Carolyn M. Noland
Notary Public, State of Alaska
Commission Expires with Office



VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130294

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130294

JUNE 30, 2009

OPERATOR'S NAME:
NOLL

OPERATOR'S NUMBER: 6520

SUBJECT'S LAST NAME:

CALIBRATION

SUBJECT'S FIRST NAME/MI :
VERIFICATION/OF

D.L. #: 012345

DEPT/AGENCY: ANCL

CASE/REPORT: 09-00000

TEST TYPE: V

ALCO TARGET VALUE: .084 ✓

ALCO S/N: X172946 ✓

--- BREATH ANALYSIS ---

.084 ADJUSTED FOR 30.16 in		
ALCO TARGET	.084	01:15
BLANK TEST	.000	01:15
INTERNAL STANDARD	VERIFIED	01:15
ALCO TV 30.16 in	.082 ✓	01:17
BLANK TEST	.000	01:18
SUBJECT SAMPLE	.000	01:19
BLANK TEST	.000	01:19
ALCO TV 30.16 in	.082 ✓	01:20
BLANK TEST	.000	01:20

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130294

JUNE 30, 2009
TIME 01:13

--- DIAGNOSTIC CHECK ---

COMPUTER:	OKAY
PROGRAM:	OKAY
SOFTWARE DATE:	01/08/09
HEATERS	
SAMPLE CHAMBER:	49c
BREATH TUBE:	42c
BAROMETER:	30.14 in
FLOW DETECTOR:	OKAY
PUMP	
HIGH SPEED:	OKAY
DETECTOR:	OKAY
FILTERS:	OKAY
QUARTZ STANDARD:	OKAY
CALIBRATION:	OKAY

PRINTER TEST

!"#\$%&'()*+,-./0123456789:;<=>?@ABCDEFGH
IJKLMNOPQRSTUVWXYZ[\]^_`abcdefghijklmnopqr
stuvwxyz{|}~Δ