


VERIFICATION OF CALIBRATION REPORT

Of DataMaster cdm Breath test Instrument
State of Alaska

Scientific Crime Detection Laboratory – Statewide Breath Alcohol Program

DataMaster cdm S/N: <u>130298</u> ✓		
Supervisor/Operator Performing the Verification Procedure:		
Name: <u>JAMES C. GIPSON</u> ID#: <u>4842</u> Date: <u>27 April 08</u> ✓		
A	Agency: <u>Palmer Police Department</u> Phone #: <u>(907) 745-4811</u>	
	Instrument Location: <u>Palmer Police Department</u> ✓	
Alco S/N: <u>X172988</u> ✓ Target Value: <u>.080</u> ✓ High Pressure: <u>1270</u>		
B	Alco Test Value Average: <u>.084</u> ✓ <u>.084</u> ✓ 1 st Alco 2 nd Alco	
Signature: 		<u>CSB</u> <u>6/13/08</u>
(OVER)		


Under the Alaska Rules of Evidence, I certify that;
(Do Not write in the area below)


(a) I am Jeanne Swartz, Forensic Scientist III, Scientific Crime Detection Laboratory, State of Alaska Department of Public Safety, and was, at the time of signing, the custodian of the records of the aforementioned laboratory and the signature was made in my official capacity as Scientific Director Designee of the State Breath Alcohol Testing Program.

(b) The verification of calibration for the above instrument has been received and reviewed. Instrument operator, Jim Gipson, whose name appears in my records is qualified to verify the calibration of the above instrument.

(c) This instrument is certified for evidentiary use and these records are being retained as original documents in my custody.

Signed and sealed,





Jeanne Swartz
Scientific Director Designee
State of Alaska
State Breath Alcohol Testing Program

6/16/08
Date

BT6 9/95

MAY 14 2008

VERIFICATION OF CALIBRATION REPORT

Of DataMaster cdm Breath test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N: 130298

(CONTINUED FROM FRONT PAGE)

C

TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130298

APRIL 27, 2008

OPERATOR'S NAME:
GIPSON/JAMES/C
OPERATOR'S NUMBER: 4842
SUBJECT'S LAST NAME:
VERIFICATION OF CAL
SUBJECT'S FIRST NAME/NI:
RJCSB

O.L. #: 0123456
DEPT/AGENCY: PLM1
CASE/REPORT: 01-2345
TEST TYPE: U
ALCO TARGET VALUE: .000 ✓
ALCO S/N: X172968 ✓

--- BREATH ANALYSIS ---

.000 ADJUSTED FOR 29.52 in
ALCO TARGET .078 14:34
BLANK TEST .000 14:35
INTERNAL STANDARD VERIFIED 14:35
ALCO TU 29.52 in .004 ✓ 14:36
BLANK TEST .000 14:36
SUBJECT SAMPLE .000 14:37
BLANK TEST .000 14:38
ALCO TU 29.54 in .004 14:38
BLANK TEST .000 14:39

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130298

APRIL 27, 2008 ✓
TIME 14:31

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY
PROGRAM: OKAY
SOFTWARE DATE: 02/20/01
HEATERS
SAMPLE CHAMBER: 49c
BREATH TUBE: 43c
BAROMETER: 29.52 in
FLOW DETECTOR: OKAY
PUMP
HIGH SPEED: OKAY
DETECTOR: OKAY
FILTERS: OKAY
QUARTZ STANDARD: OKAY
CALIBRATION: OKAY

PRINTER TEST

!"#\$%&'()*+,-./0123456789:;<=>?@ABCDEFGHIJKL
MNOPQRSTUVWXYZ[\]^_`abcdefgijklmnop
qrstuvwxyz{|}~#