

# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

FEB 20 2009

Supervisor/Operator Performing the Verification Procedure:	
Name	<u>Kyle Carrington</u> ✓
Agency	<u>University of Alaska Fairbanks Police Department</u> ✓
Instrument Location	<u>612 Yukon Drive Fairbanks, AK 99775</u>
DataMaster cdm S/N <u>130301</u> ✓	
ID#	<u>5336</u>
Date	<u>2-16-2009</u>
Alco S/N <u>x124658</u> ✓	
Target Value	<u>.082</u> ✓
High Pressure	<u>100 PSI</u>
Alco Test Values	
1 <sup>st</sup> Alco	<u>.084</u> ✓
2 <sup>nd</sup> Alco	<u>.084</u> ✓
Signature	<u>Kyle S. Carrington</u> ✓
	<u>CAB</u> <u>9/1/09</u>

(OVER)


Under the Alaska Rules of Evidence, I certify that (in the area below)

(a) I am Jeanne Swartz, Forensic Scientist III, Scientific Crime Detection Laboratory, State of Alaska Department of Public Safety, and was, at the time of signing, the custodian of the records of the aforementioned laboratory and the signature was made in my official capacity as Scientific Director Designee of the State Breath Alcohol Testing Program.

(b) The verification of calibration for the above instrument has been received and reviewed. Instrument operator, Kyle Carrington, whose name appears in my records, is qualified to verify the calibration of the above instrument.

(c) This instrument is certified for evidentiary use and these records are being retained as original documents in my custody.

Signed and sealed



Jeanne Swartz

Jeanne Swartz  
Scientific Director Designee  
State of Alaska  
State Breath Alcohol Testing Program

3/2/09

Date

# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130301 ✓

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

## NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130301

FEBRUARY 16, 2009 ✓ *COB 3/2/09*

OPERATOR'S NAME: *KC*  
CARRINGTON, KYLE  
OPERATOR'S NUMBER: 5336  
SUBJECT'S LAST NAME:  
VERIFICATION OF CAL  
SUBJECT'S FIRST NAME/MI:

O.I.L.#: 0  
DEPT/AGENCY: FBK7  
CASE/REPORT: 0  
TEST TYPE: U  
ALCO TARGET VALUE: .082 ✓  
ALCO S/N: X124658

### --- BREATH ANALYSIS ---

*COB 3/2/09*

.082 ADJUSTED FOR 29.34 in		
ALCO TARGET	.080	22:02
BLANK TEST	.000	22:03
INTERNAL STANDARD	VERIFIED	22:03
ALCO TU 29.34 in	.084 ✓	22:04
BLANK TEST	.000	22:04
SUBJECT SAMPLE	.000	22:05
BLANK TEST	.000	22:06
ALCO TU 29.34 in	.084 ✓	22:07
BLANK TEST	.000	22:07

*KC*

## DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130301

FEBRUARY 16, 2009 ✓  
TIME 21:59

### --- DIAGNOSTIC CHECK ---

COMPUTER: OKAY  
PROGRAM: OKAY  
SOFTWARE DATE: 02/20/01  
HEATERS  
SAMPLE CHAMBER: 48c  
BREATH TUBE: 43c  
BAROMETER: 29.34 in  
FLOW DETECTOR: OKAY  
PUMP  
HIGH SPEED: OKAY  
DETECTOR: OKAY  
FILTERS: OKAY  
QUARTZ STANDARD: OKAY  
CALIBRATION: OKAY

### PRINTER TEST

!"#\$%&'()\*+,-./0123456789:;<=>?@ABCDEFGHIJ  
KLMNOPQRSTUVWXYZ[\]^\_`abcdefghijklmnop  
qrstuvwxyz{|}~■