VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument
State of Alaska
Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130301

Supervisor/Operator Performing the Verification Procedure:

A

Name Kyle Carrington ID# 5336 Date 2-28-2009

Agency UAFPD Phone # 474-7727

Instrument Location 612 Yukon Drive Fairbanks, AK 99775

B

Alco S/N 77198 Target Value .075 High Pressure 1000 PSI

Alco Test Values .078 (1st Alco) .078 (2nd Alco)

Signature [signature] CB 3/10/09

(OVER)

Under the Alaska Rules of Evidence, I certify that: (Do not write in the area below).

(a) I am Jeanne Swartz, Forensic Scientist III, Scientific Crime Detection Laboratory, State of Alaska Department of Public Safety, and was, at the time of signing, the custodian of the records of the aforementioned laboratory and the signature was made in my official capacity as Scientific Director Designee of the State Breath Alcohol Testing Program.

(b) The verification of calibration for the above instrument has been received and reviewed. Instrument operator, Kyle Carrington, whose name appears in my records, is qualified to verify the calibration of the above instrument.

(c) This instrument is certified for evidentiary use and these records are being retained as original documents in my custody.

Signed and sealed,

[signature] 3/11/09

Jeanne Swartz Date
Scientific Director Designee
State of Alaska
State Breath Alcohol Testing Program

BT9 06/03

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument
State of Alaska
Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130301

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130301

FEBRUARY 28, 2009 CB 3/10/09

OPERATOR'S NAME:
CARRINGTON,KYLE
OPERATOR'S NUMBER: 5336 KC
SUBJECT'S LAST NAME:
VERIFICATION OF CAL
SUBJECT'S FIRST NAME/MI :
I
D.L. #: 0
DEPT/AGENCY: FBK7
CASE/REPORT: 0 KC
TEST TYPE: V
ALCO TARGET VALUE: .075
ALCO S/N: 77198

— BREATH ANALYSIS —

.075 ADJUSTED FOR 29.77 in

ALCO TARGET	.074	01:29
BLANK TEST	.000	01:29
INTERNAL STANDARD	VERIFIED	01:29
ALCO TV 29.77 in	.078	01:30
BLANK TEST KC	.000	01:31
SUBJECT SAMPLE CB 3/10/09	.000	01:31
BLANK TEST	.000	01:32
ALCO TV 29.77 in	.078	01:33
BLANK TEST	.000	01:33

3/11/09

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130301

FEBRUARY 28, 2009
TIME 01:40

— DIAGNOSTIC CHECK —

COMPUTER:	OKAY
PROGRAM:	OKAY
SOFTWARE DATE:	02/20/01
HEATERS	
SAMPLE CHAMBER:	48c
BREATH TUBE:	42c
BAROMETER:	29.77 in
FLOW DETECTOR:	OKAY
PUMP	
HIGH SPEED:	OKAY
DETECTOR:	OKAY
FILTERS:	OKAY
QUARTZ STANDARD:	OKAY
CALIBRATION:	OKAY

PRINTER TEST
!"#$%&'()*+,-./0123456789:;<=>?@ABCDEFG
HIJKLMNOPQRSTUVWXYZ[\]^_`abcdefghijklmno
pqrstuvwxyz{|}~