

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Supervisor/Operator Performing the Verification Procedure:		DataMaster cdm S/N	<u>130301</u>
A	Name	<u>STEVE DUNN</u>	ID# <u>4106</u> Date <u>4-3-08</u>
	Agency	<u>APD</u>	Phone # <u>786-8900</u>
	Instrument Location	<u>ANCHORAGE JAIL</u>	
B	Alco S/N	<u>X124625</u>	Target Value <u>.086</u> High Pressure <u>1250</u>
	Alco Test Value Average	<u>.081</u>	<u>.082</u>
		1 st Alco	2 nd Alco
	Signature		
		<u>CSB</u> <u>4/8/08</u>	
		(OVER)	

Under the Alaska Rules of Evidence, I certify that;

(a) I am Jeanne Swartz, Forensic Scientist III, Scientific Crime Detection Laboratory, State of Alaska Department of Public Safety, and was, at the time of signing, the custodian of the records of the aforementioned laboratory and the signature was made in my official capacity as Scientific Director Designee of the State Breath Alcohol Testing Program.

(b) The verification of calibration for the above instrument has been received and reviewed. Instrument operator, Steve P Dunn, whose name appears in my records, is qualified to verify the calibration of the above instrument.

(c) This instrument is certified for evidentiary use and these records are being retained as original documents in my custody.



Signed and sealed,



Jeanne Swartz
Scientific Director Designee
State of Alaska
State Breath Alcohol Testing Program

4/14/08

Date

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130301

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

S. DUNN #1361

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130301

APRIL 03, 2008 *COB 4/8/08*

OPERATOR'S NAME:
DUNN/S

OPERATOR'S NUMBER: 4106

SUBJECT'S LAST NAME:
VERIFICATION OF CAL

SUBJECT'S FIRST NAME/NI:
U

O.L. #: U

DEPT/AGENCY: ANC1

CASE/REPORT: U

TEST TYPE: U

ALCO TARGET VALUE: .006

ALCO S/N: X124625

--- BREATH ANALYSIS ---

.006 ADJUSTED FOR 30.09 in		
ALCO TARGET	.006	13:25
BLANK TEST	.000	13:26
INTERNAL STANDARD	VERIFIED	13:26
ALCO TV 30.09 in	.001	13:26
BLANK TEST	.000	13:27
SUBJECT SAMPLE	.000 <i>COB 4/8/08</i>	13:28
BLANK TEST	.000	13:29
ALCO TV 30.09 in	.002	13:29
BLANK TEST	.000	13:30

DIAGNOSTIC CHECK

S. DUNN #1361

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130301

APRIL 03, 2008
TIME 13:22

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY

PROGRAM: OKAY

SOFTWARE DATE: 02/20/01

HEATERS

SAMPLE CHAMBER: 49c

BREATH TUBE: 41c

BAROMETER: 30.09 in

FLOW DETECTOR: OKAY

PUMP

HIGH SPEED: OKAY

DETECTOR: OKAY

FILTERS: OKAY

QUARTZ STANDARD: OKAY

CALIBRATION: OKAY

PRINTER TEST

!"#\$%&'()*+,-./0123456789:;<=>?@ABCDEF G
HIJKL MNOPQRSTU VWXYZ[\]^_`' abcde fghi jk l m n o
pqr stuvwxyz{|}~