

VERIFICATION OF CALIBRATION REPORT
of DataMaster cdm Breath Test Instrument
State of Alaska
Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130301 ✓

Supervisor/Operator Performing the Verification Procedure:
 Name RANDALL "CLAY" FARLS ID# 6170 ✓ Date 9-21-10 ✓
 A Agency UNIVERSITY POLICE DEPT. - FBKS Phone # (907) 474-7721
 Instrument Location UNIVERSITY POLICE DEPT.

B Alco S/N X172931 ✓ Target Value .082 ✓ High Pressure 1150 psi
 Alco Test Values .084 ✓ .084 ✓
 1st Alco 2nd Alco

Signature [Handwritten Signature]

(OVER)

BMB
9/29/10

(Do Not write in the area below)



I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

[Handwritten Signature]
 Nita J. Bolz
 Scientific Director
 State Breath Alcohol Program

Subscribed and sworn before me this 8th day of DEC, 2010.

[Handwritten Signature] (Notary Seal Stamp)
 Daniel L. Nicholas
 Notary Public, State of Alaska
 Commission Expires with Office



VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130301

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130301

SEPTEMBER 21, 2010

OPERATOR'S NAME:
FARIS/CLAY
OPERATOR'S NUMBER: 6170
SUBJECT'S LAST NAME:
VERIFICATION OF CAL
SUBJECT'S FIRST NAME/MI:
UOC

O.L. #: UOC
DEPT AGENCY: FBK7
CASE/REPORT: UOC
TEST TYPE: U
ALCO TARGET VALUE: .082
ALCO S/N: X172931

--- BREATH ANALYSIS ---

.082 ADJUSTED FOR 29.54 in
ALCO TARGET .080 01:48
BLANK TEST .000 01:49
INTERNAL STANDARD VERIFIED 01:49
ALCO TV 29.54 in .084 01:49
BLANK TEST .000 01:50
SUBJECT SAMPLE .000 01:51
BLANK TEST .000 01:52
ALCO TV 29.54 in .084 01:52
BLANK TEST .000 01:53

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130301

SEPTEMBER 21, 2010
TIME 01:43

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY
PROGRAM: OKAY
SOFTWARE DATE: 02/20/01
HEATERS
SAMPLE CHAMBER: 49c
BREATH TUBE: 42c
BAROMETER: 29.54 in
FLOW DETECTOR: OKAY
PUMP
HIGH SPEED: OKAY
DETECTOR: OKAY
FILTERS: OKAY
QUARTZ STANDARD: OKAY
CALIBRATION: OKAY

PRINTER TEST

!@#%&'()*+,-./:0123456789:K=/?@#0DEF6
HIJKLWOPQRSTUWXYZ\]^_`abcde:gnijklano
pqrstuvwz{|}~