

# VERIFICATION OF CALIBRATION REPORT

MAR 29 2011

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Supervisor/Operator Performing the Verification Procedure:		DataMaster cdm S/N	<u>130302</u> ✓
Name	<u>CHRIS HILL</u>	ID#	<u>3839</u> ✓
		Date	<u>3/22/11</u> ✓
A	Agency	<u>AST</u>	Phone # <u>486-4122</u>
Instrument Location <u>KODIAK POST INTERVIEW ROOM</u>			
B	Alco S/N	<u>X124658</u> ✓	Target Value <u>.084</u> ✓
			High Pressure <u>1150</u>
	Alco Test Value Average	<u>.086</u> ✓	<u>.086</u> ✓
		1 <sup>st</sup> Alco	2 <sup>nd</sup> Alco
Signature <u>[Signature]</u>			

(OVER)

*[Handwritten initials]*

(Do Not write in the area below)

I, Nita J. Bolz, after being first duly sworn, depose and state as follows:


- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

[Signature]  
Nita J. Bolz  
Scientific Director  
State Breath Alcohol Program



Subscribed and sworn before me this 31 day of MAR, ~~2010~~ <sup>2011</sup>.

[Signature] (Notary Seal Stamp)  
Daniel L. Nicholas  
Notary Public, State of Alaska  
Commission Expires with Office



# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130302

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

## NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130302

MARCH 22, 2011

OPERATOR'S NAME: HILL C W  
OPERATOR'S NUMBER: 3835  
SUBJECT'S LAST NAME:  
VERIFICATION OF CAL  
SUBJECT'S FIRST NAME/NI :  
X  
O.L. #: X  
DEPT/AGENCY: KODG  
CASE/REPORT:  
TEST TYPE: V  
ALCO TARGET VALUE: .084  
ALCO S/N: X124658

### --- BREATH ANALYSIS ---

.084 ADJUSTED FOR 29.62 in  
ALCO TARGET .083 09:26  
BLANK TEST .000 09:27  
INTERNAL STANDARD VERIFIED 09:27  
ALCO TV 29.62 in .006 09:28  
BLANK TEST .000 09:29  
SUBJECT SAMPLE .000 09:30  
BLANK TEST .000 09:31  
ALCO TV 29.62 in .006 09:31  
BLANK TEST .000 09:32

## DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130302

MARCH 22, 2011  
TIME 09:12

### --- DIAGNOSTIC CHECK ---

COMPUTER: OKAY  
PROGRAM: OKAY  
SOFTWARE DATE: 01/08/99  
HEATERS  
SAMPLE CHAMBER: 49c  
BREATH TUBE: 42c  
BAROMETER: 29.62 in  
FLOW DETECTOR: OKAY  
PUMP  
HIGH SPEED: OKAY  
DETECTOR: OKAY  
FILTERS: OKAY  
QUARTZ STANDARD: OKAY  
CALIBRATION: OKAY

### PRINTER TEST

!@#%&'()\*+,-./:0123456789:;<=>?@BCDEFG  
HIJKLMNOPQRSTUVWXYZ[\]^\_`abcd efghijklmnop  
qrs tuvwxyz{|}~\*