

# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Supervisor/Operator Performing the Verification Procedure:		DataMaster cdm S/N <u>130802</u>
Name <u>Kevin Nelson</u>	ID# <u>5349</u>	Date <u>2/22/08</u>
A Agency <u>North Pole Police Department</u>		Phone # <u>(907)488-0612</u>
Instrument Location <u>125 Snowman Ln, North Pole, AK 99705</u>		
Alco S/N <u>X124634</u> ✓		Target Value <u>.081</u> ✓ High Pressure _____
B Alco Test Values <u>.079</u>		<u>.079</u>
1 <sup>st</sup> Alco		2 <sup>nd</sup> Alco
Signature <u>[Handwritten Signature]</u>		<u>02/22/08</u>

(OVER)


(Do Not write in the area below)

Under the Alaska Rules of Evidence, I certify that;

(a) I am Jeanne Swartz, Criminalist III, Scientific Crime Detection Laboratory, State of Alaska Department of Public Safety, and was, at the time of signing, the custodian of the records of the aforementioned laboratory and the signature was made in my official capacity as Scientific Director Designee of the State Breath Alcohol Testing Program.

(b) The verification of calibration for the above instrument has been received and reviewed. Instrument operator, Kevin Nelson, whose name appears in my records, is qualified to verify the calibration of the above instrument.

(c) This instrument is certified for evidentiary use and these records are being retained as original documents in my custody.



Signed and sealed,

[Handwritten Signature]

3/1/08

Jeanne Swartz  
Scientific Director Designee  
State of Alaska  
State Breath Alcohol Testing Program

Date

BT9 06/03

FEB 25 2008

# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N

130302

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

## NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130302

FEBRUARY 22, 2008

OPERATOR'S NAME:  
NELSON, KEVIN HOON  
OPERATOR'S NUMBER: 5349  
SUBJECT'S LAST NAME:

VOC  
SUBJECT'S FIRST NAME/MI :  
VOC

O.L. #: 00000  
DEPT/AGENCY: MPL1  
CASE/REPORT: 00-0000  
TEST TYPE: V  
ALCO TARGET VALUE: .081  
ALCO S/N: X124634

### --- BREATH ANALYSIS ---

.081 ADJUSTED FOR 29.25 in  
ALCO TARGET .079 06:40  
BLANK TEST .000 06:41  
INTERNAL STANDARD VERIFIED 06:41  
ALCO TV 29.25 in .079 06:41  
BLANK TEST .000 06:42  
SUBJECT SAMPLE .000 06:43  
BLANK TEST .000 06:44  
ALCO TV 29.25 in .079 06:44  
BLANK TEST .000 06:45

## DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130302

FEBRUARY 22, 2008  
TIME 06:37

COMPUTER:

PROGRAM:

SOFTWARE DATE: 02/20/01

HEATERS

SAMPLE CHAMBER: 4%  
BREATH TUBE: 4%

BAROMETER: 29.25 in

FLOW DETECTOR: OKAY

PUMP

HIGH SPEED: OKAY

DETECTOR: OKAY

FILTERS: OKAY

QUARTZ STANDARD: OKAY

CALIBRATION: OKAY

### PRINTER TEST

!""#%&'()\*+,-./0123456789:;<=>?@ABCDEFGHI  
HIJKLMNOPQRSTUVWXYZ[\]^\_`abcde fghijklmno  
pqrstuvwxyz{|}~