VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument
State of Alaska
Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130305

Supervisor/Operator Performing the Verification Procedure:

A

Name ANDRE ACHEE ID# 2385 Date 1/13/2009

Agency BETHEL POLICE DEPT. BOX 809 BETHEL Phone # (907)543-3781

Instrument Location BPD SQUAD ROOM

B

Alco S/N X301765 Target Value .078 High Pressure 700

Alco Test Values ~~.077~~ .078 (2/10/09 CMW) .078
1st Alco 2nd Alco

Signature [signature] CAB 2/18/09

(OVER)

Under the Alaska Rules of Evidence, I certify that, (Do Not write in the area below)

(a) I am Jeanne Swartz , Forensic Scientist III, Scientific Crime Detection Laboratory, State of Alaska Department of Public Safety, and was, at the time of signing, the custodian of the records of the aforementioned laboratory and the signature was made in my official capacity as Scientific Director Designee of the State Breath Alcohol Testing Program.

(b) The verification of calibration for the above instrument has been received and reviewed. Instrument operator, Andre Achee , whose name appears in my records, is qualified to verify the calibration of the above instrument.

(c) This instrument is certified for evidentiary use and these records are being retained as original documents in my custody.

Signed and sealed,

[signature] 2/18/09

Jeanne Swartz Date
Scientific Director Designee
State of Alaska
State Breath Alcohol Testing Program

FEB 10 2009

BT9 06/03

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument
State of Alaska
Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130305

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130305

JANUARY 13, 2009

CDB 2/18/09

OPERATOR'S NAME:
ACHEE/ANDRE/A
OPERATOR'S NUMBER: 2385
SUBJECT'S LAST NAME:
TEST
SUBJECT'S FIRST NAME/MI :
VCR
O.L. #: 1234567
DEPT/AGENCY: BET1
CASE/REPORT: XXXXXX
TEST TYPE: T
ALCO TARGET VALUE: .078
ALCO S/N: X301765

— BREATH ANALYSIS —

.078 ADJUSTED FOR 29.81 in

ALCO TARGET	.077	11:39
BLANK TEST	.000	11:40
INTERNAL STANDARD	VERIFIED	11:40
ALCO TV 29.81 in	.078	11:40
BLANK TEST	.000	11:41
SUBJECT SAMPLE	.000	11:42
BLANK TEST	.000	11:43
ALCO TV 29.83 in	.078	11:43
BLANK TEST	.000	11:44

CDB 2/18/09

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130305

JANUARY 13, 2009
TIME 11:36

— DIAGNOSTIC CHECK —

COMPUTER:	OKAY
PROGRAM:	OKAY
SOFTWARE DATE:	02/20/01
HEATERS	
SAMPLE CHAMBER:	48c
BREATH TUBE:	42c
BAROMETER:	29.81 in
FLOW DETECTOR:	OKAY
PUMP	
HIGH SPEED:	OKAY
DETECTOR:	OKAY
FILTERS:	OKAY
QUARTZ STANDARD:	OKAY
CALIBRATION:	OKAY

PRINTER TEST
!"#$%&'()*+,-./0123456789:;<=>?@ABCDEFG
HIJKLMNOPQRSTUVWXYZ[\]^_`abcdefghijklmno
pqrstuvwxyz{|}~