

# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Supervisor/Operator Performing the Verification Procedure:		DataMaster cdm S/N <u>130305</u> ✓
Name	<u>ANDRE ACHEE</u>	ID# <u>2385</u> Date <u>6/28/2008</u>
A Agency	<u>BETHEL POLICE DEPT</u>	Phone # <u>543-3781</u>
Instrument Location		<u>BETHEL P.D. SQUAD</u> ✓
Alco S/N <u>X172939</u> ✓		Target Value <u>.081</u> ✓ High Pressure <u>500</u>
B Alco Test Values	<u>.080</u> 1 <sup>st</sup> Alco	<u>.080</u> 2 <sup>nd</sup> Alco
Signature <u>[Handwritten Signature]</u> <u>ACHEE</u>		<u>083</u> <u>7/7/08</u>

(OVER)

Under the Alaska Rules of Evidence, I certify that (Do Not Write in the area below)

(a) I am Jeanne Swartz, Forensic Scientist III, Scientific Crime Detection Laboratory, State of Alaska Department of Public Safety, and was, at the time of signing, the custodian of the records of the aforementioned laboratory and the signature was made in my official capacity as Scientific Director Designee of the State Breath Alcohol Testing Program.

(b) The verification of calibration for the above instrument has been received and reviewed. Instrument operator, Andre Achee, whose name appears in my records, is qualified to verify the calibration of the above instrument.

(c) This instrument is certified for evidentiary use and these records are being retained as original documents in my custody.



Signed and sealed,

[Handwritten Signature]

Jeanne Swartz  
Scientific Director Designee  
State of Alaska  
State Breath Alcohol Testing Program

7/8/08

Date

# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130305

(CONTINUED FROM FRONT PAGE)

**C** TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

## NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130305

JUNE 28, 2008 *COB 7/1/08*

OPERATOR'S NAME: NCHEE/ANDRE  
OPERATOR'S NUMBER: 2385  
SUBJECT'S LAST NAME: UOC  
SUBJECT'S FIRST NAME/NI: UOC  
O.L. #: 1234567  
DEPT/AGENCY: BET1  
CASE/REPORT: XXXXXX  
TEST TYPE: *V8 7/1/08*  
ALCO TARGET VALUE: .081  
ALCO S/N: X172979

--- BREATH ANALYSIS ---

.081 ADJUSTED FOR 29.96 in		
ALCO TARGET	.081	06:39
BLANK TEST	.000	06:40
INTERNAL STANDARD	VERIFIED	06:40
ALCO @ 29.96 in	.080 <i>V</i>	06:41
BLANK TEST	.000 <i>V</i>	06:42
SUBJECT SAMPLE	.080 <i>COB 7/1/08</i>	06:42
BLANK TEST	.000 <i>V</i>	06:43
ALCO @ 29.96 in	.080 <i>V</i>	06:43
BLANK TEST	.000 <i>V</i>	06:44

## DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130305

JUNE 28, 2008 *✓*  
TIME 06:45

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY  
PROGRAM: OKAY  
SOFTWARE DATE: 02/20/01  
HEATERS  
SAMPLE CHAMBER: 49c  
BREATH TUBE: 43c  
BAROMETER: 29.96 in  
FLOW DETECTOR: OKAY  
PUMP  
HIGH SPEED: OKAY  
DETECTOR: OKAY  
FILTERS: OKAY  
QUARTZ STANDARD: OKAY  
CALIBRATION: OKAY

PRINTER TEST  
!"#\$%&'()\*+,-./0123456789:;<=>?@AB[DEFG  
HIJKLMNPQRSTU VWXYZ \]^\_`abcd-efghijklmno  
pqrs tuvwxyz{|}~