

VERIFICATION OF CALIBRATION REPORT

of Datamaster cdm

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

JUN 12 2009

Supervisor/Operator Performing the Verification Procedure: Datamaster cdm S/N <u>130306</u> ✓	
Name <u>PATRICK MARTIN</u>	ID# <u>4072</u> Date <u>6-12-09</u> ✓
Agency <u>APD</u>	Phone # _____
Instrument Location <u>4th Ave Sub (1st Room)</u> ✓	
Alco S/N <u>X301563</u> ✓ Target Value <u>.079</u> ✓ High Pressure <u>1100</u>	
Alco Test Value Average <u>.077</u> ✓ <u>.078</u> ✓ 1 st Alco 2 nd Alco	
Signature <u>[Handwritten Signature]</u> ✓	

CAB 6/25/09
6/25/09

(OVER)

(Do Not write in the area below)


I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.

[Handwritten Signature: Nita J. Bolz]
Nita J. Bolz
Scientific Director
State Breath Alcohol Program

Subscribed and sworn before me this 20th day of June, 2009.

[Handwritten Signature: Carolyn M. Noland]
(Notary Seal Stamp)
Carolyn M. Noland
Notary Public, State of Alaska
Commission Expires with Office



VERIFICATION OF CALIBRATION REPORT

of Datamaster cdm

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Datamaster cdm S/N 130306 ✓

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130306

JUNE 12, 2009 ✓ *08/06/09*

OPERATOR'S NAME:
MARTIN, PATRICK E.
OPERATOR'S NUMBER: 4072
SUBJECT'S LAST NAME:
UC
SUBJECT'S FIRST NAME/MI :
U
O.L. #: 1234567
DEPT/AGENCY: ANCI
CASE/REPORT: UDC
TEST TYPE: U
ALCO TARGET VALUE: .079 ✓
ALCO S/N: 1201563 ✓

— BREATH ANALYSIS —

.079 ADJUSTED FOR 29.99 in ✓
ALCO TARGET .079 09:09
BLANK TEST .000 09:10
INTERNAL STANDARD VERIFIED 09:10
ALCO TU 29.99 in .077 ✓ 09:11
BLANK TEST .000 09:11
SUBJECT SAMPLE *08/06/09* .000 09:12
BLANK TEST .000 09:13
ALCO TU 29.99 in .078 ✓ 09:13
BLANK TEST .000 09:14

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130306

JUNE 12, 2009 ✓
TIME 09:05

— DIAGNOSTIC CHECK —

COMPRESSOR: OKAY
PROGRAM: OKAY
SOFTWARE DATE: 02/20/01
HEATERS
SAMPLE CHAMBER: 49c
BREATH TUBE: 42c
BAROMETER: 29.99 in
FLOW DETECTOR: OKAY
PUMP
HIGH SPEED: OKAY
DETECTOR: OKAY
FILTERS: OKAY
QUARTZ STANDARD: OKAY
CALIBRATION: OKAY

PRINTER TEST

!@#%&'()*+,-./0123456789:;<=>?@BCDEFG
HIJKLMNOPQRSTUVWXYZ\]^_`abcdefgijklmnop
qrstuvwxyz{|}~@