

# VERIFICATION OF CALIBRATION REPORT

of Datamaster cdm

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Supervisor/Operator Performing the Verification Procedure:		Datamaster cdm S/N	130306 ✓			
Name	Patrick E. Martin	ID#	4072	Date	9-10-09 ✓	
A	Agency	APD	Phone #	788-8900		
Instrument Location		4th Ave Sub (1st Room) ✓				
B	Alco S/N	X124646 ✓	Target Value	.081 ✓	High Pressure	1300+
Alco Test Value Average		.073 ✓	.075 ✓			
		1 <sup>st</sup> Alco	2 <sup>nd</sup> Alco			
Signature		Patrick E. Martin ✓		CDB 9/28/09		

(OVER)

(Do Not write in the area below)

I, Nita J. Bolz, after being first duly sworn, depose and state as follows:


- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

*Nita J. Bolz*  
Nita J. Bolz  
Scientific Director  
State Breath Alcohol Program

Subscribed and sworn before me this 8<sup>th</sup> day of Oct, 2009.

*Carolyn M. Noland*  
(Notary Seal Stamp)

Carolyn M. Noland  
Notary Public, State of Alaska  
Commission Expires with Office



# VERIFICATION OF CALIBRATION REPORT

of Datamaster cdm

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Datamaster cdm S/N 130306 ✓

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

## NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130306

SEPTEMBER 10, 2009 ✓ *CD*

OPERATOR'S NAME: *CD*  
MARTIN, PATRICK E  
OPERATOR'S NUMBER: 4072  
SUBJECT'S LAST NAME: MARTIN  
SUBJECT'S FIRST NAME/NI: PATRICK E  
O.L. #: 1234567  
DEPT/AGENCY: ANCI  
CASE/REPORT: UOC  
TEST TYPE: U  
ALCO TARGET VALUE: .081 ✓  
ALCO S/N: X124646 ✓

### --- BREATH ANALYSIS ---

.081 ADJUSTED FOR 29.28 in		
ALCO TARGET	.079	15:00
BLANK TEST	.000	15:01
INTERNAL STANDARD	VERIFIED	15:01
ALCO TU 29.28 in	.073 ✓	15:02
BLANK TEST	.000 ✓	15:03
SUBJECT SAMPLE	.000	15:03
BLANK TEST	.000	15:04
ALCO TU 29.28 in	.075 ✓	15:06
BLANK TEST	.000	15:07

## DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130306

SEPTEMBER 10, 2009 ✓  
TIME 14:58

### --- DIAGNOSTIC CHECK ---

COMPUTER: OKAY  
PROGRAM: OKAY  
SOFTWARE DATE: 02/28/01  
HEATERS  
SAMPLE CHAMBER: 49c  
BREATH TUBE: 42c  
BAROMETER: 29.28 in  
FLOW DETECTOR: OKAY  
PUMP  
HIGH SPEED: OKAY  
DETECTOR: OKAY  
FILTERS: OKAY  
QUARTZ STANDARD: OKAY  
CALIBRATION: OKAY

### PRINTER TEST

!@#%&'()\*+,-./0123456789:;<=>?@BCDEFG  
HIJKLMNOPQRSTUVWXYZ[\]^\_`abcdefg h i j k l m n o  
p q r s t u v w x y z { | } ~