

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

MAR 25 2009

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Supervisor/Operator Performing the Verification Procedure: DataMaster cdm S/N X124635 130307 3-21-09

Name JAMES I. JOHNSON, JR ID# 3393 Date 3-21-09

A Agency YAKUTAT DPS Phone # 784-3206

Instrument Location AT POLICE STATION

B Alco S/N X124635 Target Value .079 High Pressure 900

Alco Test Values .080 - .080
1st Alco 2nd Alco

Signature [Signature] COB
4/14/09

(OVER)

(Do Not write in the area below)

Under the Alaska Rules of Evidence, I certify that;

- (a) I am Jeanne Swartz, Forensic Scientist III, Scientific Crime Detection Laboratory, State of Alaska Department of Public Safety, and was, at the time of signing, the custodian of the records of the aforementioned laboratory and the signature was made in my official capacity as Scientific Director Designee of the State Breath Alcohol Testing Program.
- (b) The verification of calibration for the above instrument has been received and reviewed. Instrument operator, James I Jensen, whose name appears in my records, is qualified to verify the calibration of the above instrument.
- (c) This instrument is certified for evidentiary use and these records are being retained as original documents in my custody.

Signed and sealed



[Signature]

Jeanne Swartz
Scientific Director Designee
State of Alaska
State Breath Alcohol Testing Program

4/16/09

Date

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N

~~X-124635~~
130307 8 4/16/09

(CONTINUED FROM FRONT PAGE)

C

TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130307

MARCH 21, 2009

OPERATOR'S NAME:

JENSEN JIM

OPERATOR'S NUMBER: 3393

SUBJECT'S LAST NAME:

UOC

SUBJECT'S FIRST NAME/MI :

UOC

O.L. #: L

DEPT/AGENCY: VAK1

CASE/REPORT: UOC

TEST TYPE: U

ALCO TARGET VALUE: .079

ALCO S/N: X124635

--- BREATH ANALYSIS ---

.079 ADJUSTED FOR 29.98 in
ALCO TARGET .079 13:02
BLANK TEST .000 13:03
INTERNAL STANDARD VERIFIED 13:03
ALCO TV 29.98 in .000 13:03
BLANK TEST .000 13:04
SUBJECT SAMPLE .000 13:05
BLANK TEST .000 13:06
ALCO TV 29.98 in .000 13:06
BLANK TEST .000 13:07

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130307

MARCH 21, 2009

TIME 12:59

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY

PROGRAM: OKAY

SOFTWARE DATE: 02/20/01

HEATERS

SAMPLE CHAMBER: 50c

BREATH TUBE: 41c

BAROMETER: 29.98 in

FLOW DETECTOR: OKAY

PUMP
HIGH SPEED: OKAY

DETECTOR: OKAY

FILTERS: OKAY

QUARTZ STANDARD: OKAY

CALIBRATION: OKAY

PRINTER TEST

!@#%&'()*+,-./0123456789:;<=>?@ABCDEF
HIJKLNPQRSTUUVWXYZ\]^_`abcde fghi jklmno
pqrstuvwxyz{|}~