

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Supervisor/Operator Performing the Verification Procedure:		DataMaster cdm S/N	130309 -			
Name	P. Nelson	ID#	4389	Date	11-2-08 ✓	
A	Agency	AST	Phone #	373-8337		
Instrument Location		MATSU WEST ✓				
B	Alco S/N	78890 ✓	Target Value	.074 ✓	High Pressure	900 ✓
Alco Test Values		.078 ✓	.078 ✓			
		1 st Alco	2 nd Alco			
Signature		P. Nelson ✓		CSB 11/2/08		


(OVER)

Under the Alaska Rules of Evidence, I certify that (Print name in the area below)

(a) I am Jeanne Swartz, Forensic Scientist III, Scientific Crime Detection Laboratory, State of Alaska Department of Public Safety, and was, at the time of signing, the custodian of the records of the aforementioned laboratory and the signature was made in my official capacity as Scientific Director Designee of the State Breath Alcohol Testing Program.

(b) The verification of calibration for the above instrument has been received and reviewed. Instrument operator, Patrick Nelson, whose name appears in my records, is qualified to verify the calibration of the above instrument.

(c) This instrument is certified for evidentiary use and these records are being retained as original documents in my custody.



Signed and sealed,

Jeanne Swartz

Jeanne Swartz
Scientific Director Designee
State of Alaska
State Breath Alcohol Testing Program

11/2/08
Date

NOV 0 2008

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130309

(CONTINUED FROM FRONT PAGE)

C

TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130309

NOVEMBER 02, 2008

OPERATOR'S NAME:

MELSON PATRICK

OPERATOR'S NUMBER: 4389

SUBJECT'S LAST NAME:

VERIFICATION OF

SUBJECT'S FIRST NAME/MI :

CALIBRATION

O.L. #: 123456

DEPT/AGENCY: PLMO

CASE/REPORT: 12-3456

TEST TYPE: U

ALCO TARGET VALUE: .074

ALCO S/N: 78890

--- BREATH ANALYSIS ---

.074 ADJUSTED FOR 29.25 in		
ALCO TARGET	.072	19:22
BLANK TEST	.000	19:24
INTERNAL STANDARD	VERIFIED	19:24
ALCO TV 29.25 in	.078	19:24
BLANK TEST	.000	19:25
SUBJECT SAMPLE	.000	19:25
BLANK TEST	.000	19:26
ALCO TV 29.25 in	.078	19:27
BLANK TEST	.000	19:27

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130309

NOVEMBER 02, 2008

TIME 19:30

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY

PROGRAM: OKAY

SOFTWARE DATE: 02/20/01

HEATERS

SAMPLE CHAMBER: 49c

BREATH TUBE: 42c

BAROMETER: 29.25 in

FLOW DETECTOR: OKAY

PUMP

HIGH SPEED: OKAY

DETECTOR: OKAY

FILTERS: OKAY

QUARTZ STANDARD: OKAY

CALIBRATION: OKAY

PRINTER TEST

!"#\$%&'()*+,-./0123456789:;<=>?@ABCDEFGHIJ
KLMNOPQRSTUVWXYZ[\]^_`abcdefgijklmnop
qrstuvwxyz{|}~