

# VERIFICATION OF CALIBRATION REPORT

MAY 13 2011

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Supervisor/Operator Performing the Verification Procedure:		DataMaster cdm S/N	130404
Name	John W. Carroll	ID#	6775 Date 04/28/2011
A Agency	Fort Greely Police Department, BIG4		Phone # (907) 873-9120
Instrument Location	Fort Greely Police Department BLDG 663 Fort Greely, AK 99731		
B Alco S/N	X172958	Target Value	.087 High Pressure 610
Alco Test Values	.086	.085	
	1 <sup>st</sup> Alco	2 <sup>nd</sup> Alco	
Signature	<i>John W. Carroll</i>		

(OVER)

*8MB  
5/13/11*

(Do Not write in the area below)

I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

*Nita J. Bolz*  
 Nita J. Bolz  
 Scientific Director  
 State Breath Alcohol Program

Subscribed and sworn before me this 31 day of MAY, 2010. *NJB 2011*

*Kyle Klopfer*  
 (Notary Seal Stamp)  
 Carolyn M. Noland  
 Notary Public, State of Alaska  
 Commission Expires with Office



# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130404

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

## NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130404

APRIL 28, 2011

OPERATOR'S NAME:

CARROLL, JOHN W

OPERATOR'S NUMBER: 6775

SUBJECT'S LAST NAME:

VOC

SUBJECT'S FIRST NAME/MI :

VOC

D.L. #: 000000

DEPT/AGENCY: FT GREELY PD B164

CASE/REPORT: 00-6000

TEST TYPE: V

ALCO TARGET VALUE: .087

ALCO S/N: X172958

— BREATH ANALYSIS —

.087 ADJUSTED FOR 28.87 in

ALCO TARGET	.083	05:44
BLANK TEST	.000	05:45
INTERNAL STANDARD	VERIFIED	05:45
ALCO TV 28.87 in	.086	05:45
BLANK TEST	.000	05:46
SUBJECT SAMPLE	.000	05:46
BLANK TEST	.000	05:47
ALCO TV 28.87 in	.085	05:48
BLANK TEST	.000	05:48

## DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130404

APRIL 28, 2011

TIME 05:41

— DIAGNOSTIC CHECK —

COMPUTER:	OKAY
PROGRAM:	OKAY
SOFTWARE DATE:	01/08/09
HEATERS	
SAMPLE CHAMBER:	48c
BREATH TUBE:	42c
BAROMETER:	28.87 in
FLOW DETECTOR:	OKAY
PUMP	
HIGH SPEED:	OKAY
DETECTOR:	OKAY
FILTERS:	OKAY
QUARTZ STANDARD:	OKAY
CALIBRATION:	OKAY

PRINTER TEST

!@#%&'()\*+,-./0123456789:;<=>?@ABCDEF  
HIJKLMNOPQRSTUVWXYZ[\]^\_`abcde fghijklmno  
pqrstuvwxyz{|}~

