

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

AUG 07 2009

Supervisor Operator Performing the Verification Procedure: DataMaster cdm S/N <u>130404</u> ✓	
Name: <u>William A. Crabtree</u>	ID# <u>6316</u> Date <u>08/05/09</u> ✓
A Agency: <u>Fort Greely Police Department</u>	Phone = <u>873-9120</u>
Instrument Location: <u>Booking Room, Bldg 663 Basement</u> ✓	
B Alco S/N <u>X301572</u> ✓	Target Value <u>.079</u> ✓ High Pressure <u>1000</u>
Alco Test Values <u>.079</u> ✓	<u>.078</u> ✓
1 st Alco	2 nd Alco
Signature: <u>Wick a Crabtree</u> ✓	
OB 8/27/09	
(OVER)	

(Do Not write in the area below.)

I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

Nita J. Bolz
Nita J. Bolz
Scientific Director
State Breath Alcohol Program

Subscribed and sworn before me this 17th day of Sept, 2009.

Carolyn M. Noland (Notary Seal Stamp)

Carolyn M. Noland
Notary Public, State of Alaska
Commission Expires with Office



VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130404 ✓

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130404

AUGUST 05, 2009

OPERATOR'S NAME:
CRABTREE/WILLIAM/A
OPERATOR'S NUMBER: 6316
SUBJECT'S LAST NAME:
VOC
SUBJECT'S FIRST NAME/MI :
VOC

O.L. #: 123456
DEPT/AGENCY: B164
CASE/REPORT: VOC
TEST TYPE: V
ALCO TARGET VALUE: .079 ✓
ALCO S/N: X301572 ✓

BREATH ANALYSIS

.079 ADJUSTED FOR 28.68 in ✓
ALCO TARGET .075 03:58
BLANK TEST .000 03:59
INTERNAL STANDARD VERIFIED 03:59
ALCO TV 28.68 in .079 03:59
BLANK TEST .000 04:00
SUBJECT SAMPLE .000 04:01
BLANK TEST .000 04:02
ALCO TV 28.68 in .078 ✓ 04:02
BLANK TEST .000 04:03

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130404

AUGUST 05, 2009 ✓
TIME 03:54

DIAGNOSTIC CHECK

COMPUTER: OKAY
PROGRAM: OKAY
SOFTWARE DATE: 02/20/01
HEATERS
SAMPLE CHAMBER: 49c
BREATH TUBE: 42c
BAROMETER: 28.68 in
FLOW DETECTOR: OKAY
PUMP
HIGH SPEED: OKAY
DETECTOR: OKAY
FILTERS: OKAY
QUARTZ STANDARD: OKAY
CALIBRATION: OKAY

PRINTER TEST

!*"#%&'()*+,-./0123456789:;<=>?@ABCDEFGHIJ
KLMNOPQRSTUVWXYZ[\]^_`abcde fghijklmno
pqrstuvwxyz{|}~