

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

JUL 07 2009

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Supervisor/Operator Performing the Verification Procedure:		DataMaster cdm S/N <u>130492</u> ✓
Name	<u>Shayne Call</u>	ID# <u>4559</u> Date <u>5-20-09</u> ✓
A Agency	<u>AST</u>	Phone # <u>733-2258</u>
Instrument Location	<u>Willow AST</u> ✓	
B Alco S/N	<u>X301581</u> ✓	Target Value <u>.082</u> ✓ High Pressure <u>1200</u>
Alco Test Value Average	<u>.084</u> ✓ 1 st Alco	<u>.083</u> ✓ 2 nd Alco
Signature	<u>[Signature]</u> CSB 7/14/09	
(OVER)		

I, Nita J. Bolz, after being first duly sworn, depose and state as follows:


(Do Not write in the area below)

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.

[Signature]
Nita J. Bolz
Scientific Director
State Breath Alcohol Program

Subscribed and sworn before me this 16th day of July 2009.

[Signature] (Notary Seal Stamp)
Carolyn M. Noland
Notary Public, State of Alaska
Commission Expires with Office



VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130492

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST



ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130492

MAY 20, 2009

OPERATOR'S NAME: *COB 7/14/09*
CALT SHAYNE M
OPERATOR'S NUMBER: 4559
SUBJECT'S LAST NAME:
VERIFICATION OF CAL
SUBJECT'S FIRST NAME/MI :
V
O.L. #: V
DEPT/AGENCY: TALO
CASE/REPORT: V
TEST TYPE: V
ALCO TARGET VALUE: .002 ✓
ALCO S/N: X301581 ✓

--- BREATH ANALYSIS ---

.002 ADJUSTED FOR 29.96 in ✓
ALCO TARGET .002 17:09
BLANK TEST .000 17:10
INTERNAL STANDARD VERIFIED 17:10
ALCO TV 29.96 in .004 ✓ 17:10
BLANK TEST .000 17:11
SUBJECT SAMPLE *COB 7/14/09* .000 17:11
BLANK TEST .000 17:12
ALCO TV 29.96 in .003 ✓ 17:13
BLANK TEST .000 17:13

DIAGNOSTIC CHECK

1200 PSI

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130492

MAY 20, 2009
TIME 17:07 ✓

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY
PROGRAM: OKAY
SOFTWARE DATE: 02/20/01
HEATERS
SAMPLE CHAMBER: 49c
BREATH TUBE: 42c
BAROMETER: 29.94 in
FLOW DETECTOR: OKAY
PUMP
HIGH SPEED: OKAY
DETECTOR: OKAY
FILTERS: OKAY
QUARTZ STANDARD: OKAY
CALIBRATION: OKAY

PRINTER TEST

! " # % & ' () * + , - . / 0 1 2 3 4 5 6 7 8 9 : ; < = > ? @ A B C D E F G
H I J K L M N O P Q R S T U V W X Y Z [\] ^ _ ` a b c d e f g h i j k l m n o
p q r s t u v w x y z { | } ~