

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Supervisor/Operator Performing the Verification Procedure:		DataMaster cdm S/N <u>130493</u> ✓
Name <u>PATRICK SMITH</u>	ID# <u>6397</u>	Date <u>24 AUG 08</u> ✓
A Agency <u>CLEAR AFS SECURITY FORCES</u>		Phone # <u>585-6313</u>
Instrument Location <u>BLDG 37 CLEAR AFS</u> ✓		
Alco S/N <u>X124655</u> ✓		Target Value <u>.084</u> ✓ High Pressure <u>1000 PSI</u>
Alco Test Values <u>.086</u> ✓		<u>.087</u> ✓
1 st Alco		2 nd Alco
Signature <u>[Signature]</u> ✓		<u>CSB</u> <u>8/29/08</u>
(OVER)		

Under the Alaska Rules of Evidence, I certify that:
(Do Not Write in the area below)

- (a) I am Jeanne Swartz, Forensic Scientist III, Scientific Crime Detection Laboratory, State of Alaska Department of Public Safety, and was, at the time of signing, the custodian of the records of the aforementioned laboratory and the signature was made in my official capacity as Scientific Director Designee of the State Breath Alcohol Testing Program.
- (b) The verification of calibration for the above instrument has been received and reviewed. Instrument operator, Patrick J Smith, whose name appears in my records, is qualified to verify the calibration of the above instrument.
- (c) This instrument is certified for evidentiary use and these records are being retained as original documents in my custody.

Signed and sealed,



[Signature] 7/2/08
Date
Jeanne Swartz
Scientific Director Designee
State of Alaska
State Breath Alcohol Testing Program

AUG 25 2008

BT9 06/03

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130493 ✓

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130493

AUGUST 24, 2008

OPERATOR'S NAME:

SMITH/PATRICK/J

OPERATOR'S NUMBER: 6397

SUBJECT'S LAST NAME:

VOC

SUBJECT'S FIRST NAME/MI :

VOC

O.L. #: 0000000

DEPT/AGENCY: CLR4

CASE/REPORT: V

TEST TYPE: V

ALCO TARGET VALUE: .004 ✓

ALCO S/N: X124655 ✓

--- BREATH ANALYSIS ---

.004 ADJUSTED FOR 28.92 in
ALCO TARGET .001 16:59
BLANK TEST .000 17:00
INTERNAL STANDARD VERIFIED 17:00
ALCO TV 28.92 in .006 ✓ 17:01
BLANK TEST .000 17:01
SUBJECT SAMPLE .000 17:02
BLANK TEST .000 17:03
ALCO TV 28.92 in .007 ✓ 17:03
BLANK TEST .000 17:04

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130493

AUGUST 24, 2008 ✓

TIME 16:57

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY

PROGRAM: OKAY

SOFTWARE DATE: 02/20/01

HEATERS

SAMPLE CHAMBER: 49c

BREATH TUBE: 41c

BAROMETER: 28.92 in

FLOW DETECTOR: OKAY

PUMP

HIGH SPEED: OKAY

DETECTOR: OKAY

FILTERS: OKAY

QUARTZ STANDARD: OKAY

CALIBRATION: OKAY

PRINTER TEST

! " # \$ % & ' () * + , - . / 0 1 2 3 4 5 6 7 8 9 : ; < = > ? @ A B C D E F G
H I J K L M N O P Q R S T U V W X Y Z [\] ^ _ ` a b c d e f g h i j k l m n o
p q r s t u v w x y z { | } ~