

FEB 14 2010
2011
3/14/11

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument
State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Supervisor/Operator Performing the Verification Procedure:		DataMaster cdm S/N	130493 ✓
Name	PATRICK SMITH	ID#	6397 ✓
		Date	28 JAN 11 ✓
A	Agency	CLEAR SECURITY FORCES	Phone # 585-6313
	Instrument Location	BLDG 37 CLEAR AFS, AK	99704
B	Alco S/N	X124609 ✓	Target Value .081 ✓
			High Pressure 1100
	Alco Test Values	.075 ✓	.078 ✓
		1st Alco	2nd Alco
	Signature		


(OVER)

BMB
3/16/11

(Do Not write in the area below)

I, Nita J. Bolz, after being first duly sworn, depose and state as follows:


- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.


 Nita J. Bolz
 Scientific Director
 State Breath Alcohol Program

Subscribed and sworn before me this 14th day of MAR, 2010


(Notary Seal Stamp)

Daniel L. Nicholas
Notary Public, State of Alaska
Commission Expires with Office



VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument
State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130493

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130493

JANUARY 28, 2011

OPERATOR'S NAME:

SMITH/PATRICK/J

OPERATOR'S NUMBER: 6397

SUBJECT'S LAST NAME:

VOC

SUBJECT'S FIRST NAME/MI :

VOC

O.L. #: 1234567

DEPT/AGENCY: CLR4

CASE/REPORT: 1234567

TEST TYPE: V

ALCO TARGET VALUE: .001

ALCO S/N: X124609

--- BREATH ANALYSIS ---

.001 ADJUSTED FOR 29.70 in
ALCO TARGET .000 18:38
BLANK TEST .000 18:40
INTERNAL STANDARD VERIFIED 18:40
ALCO TV 29.70 in .075 18:40
BLANK TEST .000 18:41
SUBJECT SAMPLE .000 18:41
BLANK TEST .000 18:42
ALCO TV 29.70 in .078 18:42
BLANK TEST .000 18:43

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130493

JANUARY 28, 2011

TIME 18:18

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY
PROGRAM: OKAY
SOFTWARE DATE: 01/08/09
HEATERS
SAMPLE CHAMBER: 49c
BREATH TUBE: 41c
BAROMETER: 29.70 in
FLOW DETECTOR: OKAY
PUMP
HIGH SPEED: OKAY
DETECTOR: OKAY
FILTERS: OKAY
QUARTZ STANDARD: OKAY
CALIBRATION: OKAY

PRINTER TEST

!""#%&'()*+,-./0123456789:;<=>?@ABCDEF6
HIJKLMNOPQRSTUVWXYZ\]^_`abcde fghijklmno
pqrstuvwxyz{|}~