

# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

MAY 12 2009

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Supervisor/Operator Performing the Verification Procedure:		DataMaster cdm S/N	<u>130526</u> ✓
Name		<u>JOSEPH V. SHOENAKER</u>	ID# <u>4188</u> Date <u>5-7-09</u>
A	Agency	<u>SOP1 SAND POINT P.D.</u>	Phone# <u>(907) 383-3700</u>
Instrument Location		<u>SAND POINT, AK 99661</u>	
Alco S/N		<u>77621</u> ✓	Target Value <u>.084</u> High Pressure <u>525</u>
Alco Test Value Average		<u>.082</u> ✓ 1 <sup>st</sup> Alco	<u>.083</u> ✓ 2 <sup>nd</sup> Alco
Signature		<u>Joseph V. Shoemaker</u>	<u>COB</u> <u>5/13/09</u>

(OVER)

(Do Not write in the area below)

I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.

Nita J. Bolz  
Nita J. Bolz  
Scientific Director  
State Breath Alcohol Program

Subscribed and sworn before me this 19<sup>th</sup> day of May, 2009.

Carolyn M. Noland  
(Notary Seal Stamp)  
Carolyn M. Noland  
Notary Public, State of Alaska  
Commission Expires with Office



