

# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

APR 01 2011

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Supervisor/Operator Performing the Verification Procedure:	
DataMaster cdm S/N <u>130526</u>	
Name	<u>Joseph V. Steemaker</u> ID# <u>4188</u> Date <u>3-26-11</u>
A Agency	<u>(SPR) Sand Point</u> Phone # <u>907-383-3700</u>
Instrument Location <u>SAND POINT, AK 99661</u>	
B Alco S/N <u>X123004</u> Target Value <u>.079</u> High Pressure <u>300</u>	
Alco Test Value Average <u>.077</u> <u>.077</u>	
1 <sup>st</sup> Alco 2 <sup>nd</sup> Alco	
Signature <u>Joseph V. Steemaker</u>	
(OVER)	

*JMB*  
4/1/11

(Do Not write in the area below)


I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

Nita J. Bolz  
Nita J. Bolz  
Scientific Director  
State Breath Alcohol Program

Subscribed and sworn before me this 1<sup>st</sup> day of APRIL, 2011 *an*

Daniel L. Nicholas  
(Notary Seal Stamp)  
Daniel L. Nicholas  
Notary Public, State of Alaska  
Commission Expires with Office



# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130526

(CONTINUED FROM FRONT PAGE)

**C** TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

## NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130526

MARCH 26, 2011

OPERATOR'S NAME: SHOENAKER J V  
OPERATOR'S NUMBER: 4125  
SUBJECT'S LAST NAME: F  
SUBJECT'S FIRST NAME/MI: 6  
O.L. #: H  
DEPT/AGENCY: SDP1  
CASE/REPORT: I  
TEST TYPE: V  
ALCO TARGET VALUE: .079  
ALCO S/N: X173034

### --- BREATH ANALYSIS ---

.079 ADJUSTED FOR 29.96 in  
ALCO TARGET .079 22:23  
BLANK TEST .000 22:24  
INTERNAL STANDARD VERIFIED 22:24  
ALCO TV 29.96 in .077 22:24  
BLANK TEST .000 22:25  
SUBJECT SAMPLE .000 22:26  
BLANK TEST .000 22:27  
ALCO TV 29.96 in .077 22:27  
BLANK TEST .000 22:28

## DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130526

MARCH 26, 2011  
TIME 22:29

### --- DIAGNOSTIC CHECK ---

COMPUTER: OKAY  
PROGRAM: OKAY  
SOFTWARE DATE: 02/20/01  
HEATERS  
SAMPLE CHAMBER: 30c  
BREATH TUBE: 41c  
BAROMETER: 29.96 in  
FLOW DETECTOR: OKAY  
PUMP  
HIGH SPEED: OKAY  
DETECTOR: OKAY  
FILTERS: OKAY  
QUARTZ STANDARD: OKAY  
CALIBRATION: OKAY

### PRINTER TEST

!@#%&'()\*+,-./0123456789:;<=>?@ABCDEF  
GHIJKLMNOPQRSTUVWXYZ[\]^\_`abcde fghijklmno  
pqrstuvwxy z{|}~"