VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument
State of Alaska
Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130570

Supervisor/Operator Performing the Verification Procedure:

A

Name C. RITALA /1349 ID# 3791 Date 7/28/08

Agency ANCHORAGE POLICE DEPT. Phone # 786-2640

Instrument Location FLORA SUBSTATION (4th AVENUE)

B

Alco S/N X301554 Target Value .075 High Pressure 1200

Alco Test Values .074 (1st Alco) .074 (2nd Alco)

Signature C. Rita /1349 CRB 8/12/08

(OVER)

Under the Alaska Rules of Evidence, I certify that (Write in the area below)

(a) I am Jeanne Swartz , Forensic Scientist III, Scientific Crime Detection Laboratory, State of Alaska Department of Public Safety, and was, at the time of signing, the custodian of the records of the aforementioned laboratory and the signature was made in my official capacity as Scientific Director Designee of the State Breath Alcohol Testing Program.

(b) The verification of calibration for the above instrument has been received and reviewed. Instrument operator, Chris Ritala , whose name appears in my records, is qualified to verify the calibration of the above instrument.

(c) This instrument is certified for evidentiary use and these records are being retained as original documents in my custody.

STATE OF ALASKA SCIENTIFIC CRIME DETECTION LABORATORY

Signed and sealed,

8/13/08

Jeanne Swartz
Scientific Director Designee
State of Alaska
State Breath Alcohol Testing Program

Date

BT9 03/00

AUG 0 1 2008

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument
State of Alaska
Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130570 (Cbr)

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130570

JULY 28, 2008 ✓ CAB 8/12/08

OPERATOR'S NAME:
RITALA
OPERATOR'S NUMBER: 3791
SUBJECT'S LAST NAME:
VERIFICATION OF CALIBRATION
SUBJECT'S FIRST NAME/MI :
A
O.L. #: A
DEPT/AGENCY: ANCI
CASE/REPORT: A
TEST TYPE: V
ALCO TARGET VALUE: .075 ✓
ALCO S/N: X301554 ✓

--- BREATH ANALYSIS ---

.075 ADJUSTED FOR 29.80 in

ALCO TARGET	.074	16:44
BLANK TEST	.000	16:45
INTERNAL STANDARD	VERIFIED	16:45
ALCO TV 29.80 in	.074 ✓	16:46
BLANK TEST	.000	16:46
SUBJECT SAMPLE	.000 CAB 8/12/08	16:47
BLANK TEST	.000	16:48
ALCO TV 29.80 in	.074 ✓	16:48
BLANK TEST	.000	16:49

CR/1349 7/28/08

8/13/08

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130570

JULY 28, 2008 ✓
TIME 16:50

--- DIAGNOSTIC CHECK ---

COMPUTER:	OKAY
PROGRAM:	OKAY
SOFTWARE DATE:	03/20/01
HEATERS	
SAMPLE CHAMBER:	51c
BREATH TUBE:	42c
BAROMETER:	29.80 in
FLOW DETECTOR:	OKAY
PUMP	
HIGH SPEED:	OKAY
DETECTOR:	OKAY
FILTERS:	OKAY
QUARTZ STANDARD:	OKAY
CALIBRATION:	OKAY

PRINTER TEST
!"#$%&'()*+,-./0123456789:;<=>?@ABCDEFG
HIJKLMNOPQRSTUVWXYZ[\]^_`abcdefghijklmno
pqrstuvwxyz{|}+~

CR/1349 7/28/08