

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program


Supervisor/Operator Performing the Verification Procedure:		DataMaster cdm S/N	<u>130570</u> ✓
Name		<u>J Swartz</u>	ID# <u>40314</u> Date <u>8/18/08</u> ✓
A	Agency	<u>SCDL</u> ✓	Phone # <u>269-5592</u>
Instrument Location		<u>SCDL</u>	
Alco S/N		<u>SC0707</u> ✓	Target Value <u>.080</u> ✓ High Pressure <u>300 psi</u>
Alco Test Values		<u>.076</u> ✓ 1 st Alco	<u>.076</u> ✓ 2 nd Alco
Signature		<u>[Signature]</u>	<u>CSB</u> <u>8/21/08</u>
(OVER)			

Under the Alaska Rules of Evidence, I certify that, (Do not write in the area below)

(a) I am Jeanne Swartz, Forensic Scientist III, Scientific Crime Detection Laboratory, State of Alaska Department of Public Safety, and was, at the time of signing, the custodian of the records of the aforementioned laboratory and the signature was made in my official capacity as Scientific Director Designee of the State Breath Alcohol Testing Program.

(b) The verification of calibration for the above instrument has been received and reviewed. Instrument operator, Jeanne M Swartz, whose name appears in my records, is qualified to verify the calibration of the above instrument.

(c) This instrument is certified for evidentiary use and these records are being retained as original documents in my custody.



Signed and sealed,

[Signature]

Jeanne Swartz
Scientific Director Designee
State of Alaska
State Breath Alcohol Testing Program

8/25/08
Date

AUG 21 2008

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130570 ✓

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130570

AUGUST 18, 2008

OPERATOR'S NAME:

SWARTZ/J

OPERATOR'S NUMBER: 4314

SUBJECT'S LAST NAME:

VERIFICATION OF CAL

SUBJECT'S FIRST NAME/MI :

L

O.L. #: L

DEPT/AGENCY: ARA1

CASE/REPORT:

TEST TYPE: V

ALCO TARGET VALUE: .000 ✓

ALCO S/N: SCOTT07 ✓

--- BREATH ANALYSIS ---

.000 ADJUSTED FOR 29.49 in
ALCO TARGET .078 16:59
BLANK TEST .000 17:00
INTERNAL STANDARD VERIFIED 17:00
ALCO TV 29.49 in .076 ✓ 17:00
BLANK TEST .000 17:01
SUBJECT SAMPLE ~~.000~~ ✓ 17:02
BLANK TEST .000 17:03
ALCO TV 29.51 in .076 ✓ 17:03
BLANK TEST .000 17:04

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130570

AUGUST 18, 2008 ✓

TIME 16:57

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY
PROGRAM: OKAY
SOFTWARE DATE: 02/20/01
HEATERS
SAMPLE CHAMBER: 50c
BREATH TUBE: 42c
BAROMETER: 29.51 in
FLOW DETECTOR: OKAY
PUMP
HIGH SPEED: OKAY
DETECTOR: OKAY
FILTERS: OKAY
QUARTZ STANDARD: OKAY
CALIBRATION: OKAY

PRINTER TEST

!"#\$%&'()*+,-./0123456789:;<=>?@ABCDEF G
HIJKL MNOPQRSTUVWXYZ \]^_` abcde fghi jk lmnop
qrstuvwxyz{|}~