

# VERIFICATION OF CALIBRATION REPORT

NOV 09 2009

of DataMaster cdm Breath Test Instrument

11-8-09

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Supervisor/Operator Performing the Verification Procedure:		
DataMaster cdm S/N <u>130570</u> ✓		
Name <u>Derek Site</u>	ID# <u>4096</u> Date <u>11-8-09</u> ✓	
Agency <u>Anchorage Police Department</u>	Phone # <u>782-8900</u>	
Instrument Location <u>Flora Substation</u> ✓		
Alco S/N <u>78895</u> ✓	Target Value <u>.024</u> High Pressure <u>~ 0 psi</u> ✓	
Alco Test Values	<u>.079</u> ✓ 1 <sup>st</sup> Alco	<u>.079</u> ✓ 2 <sup>nd</sup> Alco
Signature <u>Derek Site</u> ✓		
(OVER)		

DMS  
11/14/09

(Do Not write in the area below)

I, Nita J. Bolz, after being first duly sworn, depose and state as follows:


- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

Nita J. Bolz  
Nita J. Bolz  
Scientific Director  
State Breath Alcohol Program

Subscribed and sworn before me this 24<sup>th</sup> day of November 2009.

Carolyn M. Noland (Notary Seal Stamp)

Carolyn M. Noland  
Notary Public, State of Alaska  
Commission Expires with Office



# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130570

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

## NONDRINKING SUBJECT TEST

## DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130570

NOVEMBER 08, 2009

OPERATOR'S NAME:

SITZ, DEREK W

OPERATOR'S NUMBER: 4096

SUBJECT'S LAST NAME:

VOC

SUBJECT'S FIRST NAME/MI :

VOC

O.L. #: 1234567

DEPT/AGENCY: ANCL

CASE/REPORT: 00-1234

TEST TYPE: V

ALCO TARGET VALUE: .084

ALCO S/N: 78895

BMB  
11/14/09

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130570

NOVEMBER 08, 2009

TIME 01:55

--- DIAGNOSTIC CHECK ---

COMPUTER:

OKAY

PROGRAM:

OKAY

SOFTWARE DATE:

12/11/08

HEATERS

SAMPLE CHAMBER:

50c

BREATH TUBE:

42c

BAROMETER:

29.39 in

FLOW DETECTOR:

OKAY

PUMP

HIGH SPEED:

OKAY

DETECTOR:

OKAY

FILTERS:

OKAY

QUARTZ STANDARD:

OKAY

CALIBRATION:

OKAY

DS  
1405  
11/8/09

--- BREATH ANALYSIS ---

.084 ADJUSTED FOR 29.39 in

ALCO TARGET

.082

01:51

BLANK TEST

.000

01:52

INTERNAL STANDARD

VERIFIED

01:52

ALCO TV 29.39 in

.079

01:52

BLANK TEST

.000

01:53

SUBJECT SAMPLE

.000

01:53

BLANK TEST

.000

01:54

ALCO TV 29.39 in

.079

01:54

BLANK TEST

.000

01:55

BMB  
11/14/09

PRINTER TEST

!"#\$%&'()\*+,-./0123456789:;<=>?@ABCDEFGH  
IJKLMNOPQRSTUVWXYZ[\]^\_`abcdefghijklmnopqr  
stuvwxyz{|}~Δ