

VERIFICATION OF CALIBRATION REPORT
of DataMaster cdm Breath Test Instrument
State of Alaska
Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Supervisor/Operator Performing the Verification Procedure:		DataMaster cdm S/N <u>130571</u>
Name	<u>Steve Dunn</u>	<div style="border: 1px solid black; border-radius: 50%; padding: 2px; display: inline-block;"> EMP # 4100 ID# 1367 </div> Date <u>11/10/08</u>
A Agency	<u>APD</u>	Phone # <u>786-8900</u>
Instrument Location	<u>Anchorage Jail</u> ✓	
B Alco S/N	<u>58428</u> ✓	Target Value <u>0.069</u> High Pressure <u>1250 PSI</u>
Alco Test Values	<u>0.071</u> ✓ 1 st Alco	<u>0.071</u> ✓ 2 nd Alco
Signature		
(OVER)		

Under the Alaska Rules of Evidence, I certify that (Do not write in the area below)

- (a) I am Jeanne Swartz, Forensic Scientist III, Scientific Crime Detection Laboratory, State of Alaska Department of Public Safety, and was, at the time of signing, the custodian of the records of the aforementioned laboratory and the signature was made in my official capacity as Scientific Director Designee of the State Breath Alcohol Testing Program.
- (b) The verification of calibration for the above instrument has been received and reviewed. Instrument operator, Steve P Dunn, whose name appears in my records, is qualified to verify the calibration of the above instrument.
- (c) This instrument is certified for evidentiary use and these records are being retained as original documents in my custody.



Signed and sealed,

11/21/08

Jeanne Swartz
 Scientific Director Designee
 State of Alaska
 State Breath Alcohol Testing Program

Date

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130571 ✓

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

S. DUNN #1366
made in my official capacity at
time of signing, the custodian
of the Alaska Department of Public Safety
Scientific Crime Detection Laboratory
DataMaster cdm 130571

NOVEMBER 10, 2000 ✓

OPERATOR'S NAME:
DUNN/S

OPERATOR'S NUMBER: 4100

SUBJECT'S LAST NAME:
VERIFICATION OF CAL

SUBJECT'S FIRST NAME/MI:
V

O.L. #: V

DEPT/AGENCY: ANCI

CASE/REPORT: V

TEST TYPE: V

ALCO TARGET VALUE: .069 ✓

ALCO S/N: 58428 ✓

--- BREATH ANALYSIS ---

.069 ADJUSTED FOR 29.34 in		
ALCO TARGET	.067	13:23
BLANK TEST	.000	13:24
INTERNAL STANDARD	VERIFIED	13:24
ALCO TV 29.34 in	.071 ✓	13:25
BLANK TEST	.000	13:26
SUBJECT SAMPLE	.000	13:26
BLANK TEST	.000	13:27
ALCO TV 29.34 in	.071 ✓	13:27
BLANK TEST	.000	13:28

DIAGNOSTIC CHECK

S. DUNN #1366
Department of Public Safety, Forensic
Science Laboratory and the
Department of Public Safety,
State of Alaska
November 10, 2000
TIME 13:18
The Alaska Rules of Evidence,
--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY

PROGRAM: OKAY

SOFTWARE DATE: 02/20/01

HEATERS

SAMPLE CHAMBER: 51c

BREATH TUBE: 42c

BAROMETER: 29.34 in

FLOW DETECTOR: OKAY

PUMP

HIGH SPEED: OKAY

DETECTOR: OKAY

FILTERS: OKAY

QUARTZ STANDARD: OKAY

CALIBRATION: OKAY

PRINTER TEST

"###"()*,./0123456789:;<=>?@ABCDEF
GHIJKLMNOPQRSTUVWXYZ[\]^_`abcde fghijklmno
pqrstuvwxyz{|}~"