

# VERIFICATION OF CALIBRATION REPORT

of Datamaster cdm

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Datamaster cdm S/N 130578

Supervisor/Operator Performing the Verification Procedure:

Name Michael A. Wery ID# 3590 Date 2-23-10

**A** Agency AST Phone # 451-2950

Instrument Location Cantwell

---

**B** Alco S/N X172969 Target Value .085 High Pressure 1250

Alco Test Value Average .075 .076

1<sup>st</sup> Alco 2<sup>nd</sup> Alco

Signature Michael A. Wery

(OVER)

*AMB*  
*3/2/10*


(Do Not write in the area below)

- I, Nita J. Bolz, after being first duly sworn, depose and state as follows:
- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
  - (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
  - (3) I am the Scientific Director of the State Breath Alcohol Program.
  - (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
  - (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
  - (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

Nita J. Bolz  
 Nita J. Bolz  
 Scientific Director  
 State Breath Alcohol Program

Subscribed and sworn before me this 23<sup>rd</sup> day of March 2010.

Carolyn M. Noland (Notary Seal Stamp)  
 Carolyn M. Noland  
 Notary Public, State of Alaska  
 Commission Expires with Office



# VERIFICATION OF CALIBRATION REPORT

of Datamaster cdm

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Datamaster cdm S/N 130578 ✓

(CONTINUED FROM FRONT PAGE)

C

**TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.**

## NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130578

FEBRUARY 23, 2010

OPERATOR'S NAME:

WERY, MICHAEL/A

OPERATOR'S NUMBER: 3598

SUBJECT'S LAST NAME:

VERIFICATION

SUBJECT'S FIRST NAME/MI :

OF CALIBRATION

O.L. #: 0

DEPT/AGENCY: CNT0

CASE/REPORT:

TEST TYPE: V

ALCO TARGET VALUE: .065 ✓

ALCO S/N: X172969 ✓

--- BREATH ANALYSIS ---

.065 ADJUSTED FOR 27.62 in ✓  
ALCO TARGET .078 14:46  
BLANK TEST .000 14:47  
INTERNAL STANDARD VERIFIED 14:47  
ALCO TV 27.62 in .075 ✓ 14:47  
BLANK TEST .000 14:48  
SUBJECT SAMPLE .000 14:49  
BLANK TEST .000 14:50  
ALCO TV 27.62 in .076 ✓ 14:50  
BLANK TEST .000 14:51

## DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130578 ✓

FEBRUARY 23, 2010 ✓

TIME 14:43

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY  
PROGRAM: OKAY  
SOFTWARE DATE: 02/20/01  
HEATERS  
SAMPLE CHAMBER: 58c  
BREATH TUBE: 41c  
BAROMETER: 27.61 in  
FLOW DETECTOR: OKAY  
PUMP  
HIGH SPEED: OKAY  
DETECTOR: OKAY  
FILTERS: OKAY  
QUARTZ STANDARD: OKAY  
CALIBRATION: OKAY

PRINTER TEST

!\*"#%&'()\*+,-./0123456789:;<=>@ABCDEFGHIJ  
KLMNOPQRSTUVWXYZ[\]^\_`abcde fghijklmno  
pqrstuvwxyz{|}~