VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument
State of Alaska
Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130578

Supervisor/Operator Performing the Verification Procedure:

Name Malik Jones ID# 5830 Date 5/14/08

A

Agency AST Phone # 768-2202

Instrument Location Centrex II

Alco S/N 70729 Target Value .084 High Pressure 100 psi

B

Alco Test Value Average .073 (1st Alco) .073 (2nd Alco)

Signature CDB 6/3/08

(OVER)

Under the Alaska Rules of Evidence, I certify that:
(Do Not write in the area below)

(a) I am Jeanne Swartz, Forensic Scientist III, Scientific Crime Detection Laboratory, State of Alaska Department of Public Safety, and was, at the time of signing, the custodian of the records of the aforementioned laboratory and the signature was made in my official capacity as Scientific Director Designee of the State Breath Alcohol Testing Program.

(b) The verification of calibration for the above instrument has been received and reviewed. Instrument operator, Malik Jones, whose name appears in my records, is qualified to verify the calibration of the above instrument.

(c) This instrument is certified for evidentiary use and these records are being retained as original documents in my custody.

Signed and sealed,

6/5/08

Jeanne Swartz
Scientific Director Designee
State of Alaska
State Breath Alcohol Testing Program

Date

BT9 03/00

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument
State of Alaska
Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130578

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130578

MAY 14, 2008 CDB 6/3/08

OPERATOR'S NAME:
JONES/MALIK/JAMAL
OPERATOR'S NUMBER: 5830
SUBJECT'S LAST NAME:
VERIFICATION OF CAL
SUBJECT'S FIRST NAME/MI :
0
O.L. #: 0
DEPT/AGENCY: CNT0
CASE/REPORT:
TEST TYPE: V
ALCO TARGET VALUE: .084
ALCO S/N: 70729

--- BREATH ANALYSIS ---

.084 ADJUSTED FOR 27.20 in

ALCO TARGET	.076	15:27
BLANK TEST	.000	15:28
INTERNAL STANDARD	VERIFIED	15:28
ALCO TV 27.20 in	.073	15:28
BLANK TEST	.000	15:29
SUBJECT SAMPLE	.000 CDB	15:30
BLANK TEST	.000 6/3/08	15:31
ALCO TV 27.20 in	.073	15:31
BLANK TEST	.000	15:32

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130578

MAY 14, 2008
TIME 15:33

--- DIAGNOSTIC CHECK ---

COMPUTER:	OKAY
PROGRAM:	OKAY
SOFTWARE DATE:	02/20/01
HEATERS	
SAMPLE CHAMBER:	51c
BREATH TUBE:	41c
BAROMETER:	27.20 in
FLOW DETECTOR:	OKAY
PUMP	
HIGH SPEED:	OKAY
DETECTOR:	OKAY
FILTERS:	OKAY
QUARTZ STANDARD:	OKAY
CALIBRATION:	OKAY

PRINTER TEST
!"#$%&'()*+,-./0123456789:;<=>?@ABCDEFG
HIJKLMNOPQRSTUVWXYZ[\]^_`abcdefghijklmno
pqrstuvwxyz{|}~